

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082321\
 Data File : VU044510.D
 Acq On : 23 Aug 2021 19:07
 Operator : SY/MD
 Sample : M3517-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AG7

Quant Time: Aug 24 06:18:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 24 06:16:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.321	114	110986	5.000	ug/L	0.06
28) Chlorobenzene-d5	9.439	117	109560	5.000	ug/L	0.02
58) 1,4-Dichlorobenzene-d4	11.822	152	50309	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.655	65	39696	4.422	ug/L	0.05
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.977	69	36834	4.722	ug/L	0.06
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.636	65	19341	5.065	ug/L	0.06
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.400%
20) 2-Butanone-d5	4.784	46	132485	53.368	ug/L	0.14
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.740%
24) Chloroform-d	5.179	84	76602	4.826	ug/L	0.11
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.600%
26) 1,2-Dichloroethane-d4	5.796	65	43539	5.030	ug/L	0.08
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.600%
32) Benzene-d6	5.819	84	153273	4.968	ug/L	0.08
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
36) 1,2-Dichloropropane-d6	6.755	67	50360	5.102	ug/L	0.05
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.935	98	127416	4.658	ug/L	0.03
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
43) trans-1,3-Dichloroprop...	8.214	79	15934	4.495	ug/L	0.03
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.664	63	113216	59.066	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.140%
56) 1,1,2,2-Tetrachloroeth...	10.767	84	42608	5.126	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.600%
66) 1,2-Dichlorobenzene-d4	12.201	152	45400	5.341	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082321\
Data File : VU044510.D
Acq On : 23 Aug 2021 19:07
Operator : SY/MD
Sample : M3517-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AG7

Quant Time: Aug 24 06:18:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 24 06:16:49 2021
Response via : Initial Calibration

