

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU082418\
 Data File : VU026351.D
 Acq On : 24 Aug 2018 22:09
 Operator : MD/SY
 Sample : J4644-19 40X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETOM9

Quant Time: Aug 25 05:03:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 25 04:25:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	291693	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	283548	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	136076	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	72721	40.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.74%
7) Chloroethane-d5	1.68	69	59123	40.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.50%
11) 1,1-Dichloroethene-d2	2.27	63	109538	31.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.64%
21) 2-Butanone-d5	4.18	46	138919	97.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.83%
24) Chloroform-d	4.65	84	171637	43.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.56%
26) 1,2-Dichloroethane-d4	5.31	65	118976	46.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.20%
32) Benzene-d6	5.35	84	336378	44.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.68%
36) 1,2-Dichloropropane-d6	6.33	67	111106	45.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.24%
41) Toluene-d8	7.57	98	300559	41.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.90%
43) trans-1,3-Dichloropropene-	7.85	79	52312	41.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.14%
47) 2-Hexanone-d5	8.31	63	92857	89.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.20%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	161352	44.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	133276	46.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.32%

Target Compounds


					Qvalue
5) Vinyl chloride	1.40	62	703585	330.693 ug/L	96
17) trans-1,2-Dichloroethene	2.99	96	52009	27.484 ug/L	95
20) cis-1,2-Dichloroethene	4.23	96	3604867	1725.826 ug/L	97
34) Trichloroethene	6.19	95	4288	2.098 ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU082418\
Data File : VU026351.D
Acq On : 24 Aug 2018 22:09
Operator : MD/SY
Sample : J4644-19 40X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETOM9

Quant Time: Aug 25 05:03:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 25 04:25:01 2018
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 330.693 ug/L

RT: 1.40 min Scan# 252

Delta R.T. 0.00 min

Lab File: VU026351.D

Acq: 24 Aug 2018 22:09

Instrument :

MSVOA_U

ClientSampleId :





ETOM9

Tgt Ion: 62 Resp: 703585

Ion Ratio Lower Upper

62 100

64 32.1 24.2 45.0



#17

trans-1,2-Dichloroethene

Concen: 27.484 ug/L

RT: 2.99 min Scan# 745

Delta R.T. 0.00 min

Lab File: VU026351.D

Acq: 24 Aug 2018 22:09




Tgt Ion: 96 Resp: 52009


Ion Ratio Lower Upper

96 100

61 145.1 97.9 181.9

98 70.2 45.1 83.7





#20

cis-1,2-Dichloroethene

Concen: 1725.826 ug/L

RT: 4.23 min Scan# 1132

Delta R.T. 0.00 min

Lab File: VU026351.D

Acq: 24 Aug 2018 22:09

Instrument :

MSVOA_U

ClientSampled :

ETOM9





Tgt Ion: 96 Resp: 3604867

Ion Ratio Lower Upper

96 100

61 125.9 90.3 167.7

68 0.0 0.0 0.0



#34

Trichloroethene

Concen: 2.098 ug/L

RT: 6.19 min Scan# 1741

Delta R.T. 0.00 min

Lab File: VU026351.D

Acq: 24 Aug 2018 22:09

Tgt Ion: 95 Resp: 4288

Ion Ratio Lower Upper

95 100

97 72.6 47.2 87.6

132 105.2 70.2 130.4

130 110.0 73.4 136.4

