

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082420\
 Data File : VU039949.D
 Acq On : 24 Aug 2020 16:11
 Operator : SY/MD
 Sample : L3767-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFG5

Quant Time: Aug 25 02:17:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 25 02:02:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	71231	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	73155	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	34937	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	24891	4.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.00%
7) Chloroethane-d5	1.92	69	22316	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.60%
11) 1,1-Dichloroethene-d2	2.58	63	41473	3.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.00%
20) 2-Butanone-d5	4.66	46	82937	41.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.42%
24) Chloroform-d	5.08	84	48409	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.72	65	28274	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.00%
32) Benzene-d6	5.75	84	94052	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.40%
36) 1,2-Dichloropropane-d6	6.70	67	31689	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.20%
41) Toluene-d8	7.91	98	84177	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
43) trans-1,3-Dichloropropene-	8.19	79	12328	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
46) 2-Hexanone-d5	8.64	63	61038	39.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.80%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	24566	3.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	34343	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds					Ovalue
3) Chloromethane	1.53	50	4685	0.559 ug/L	93
13) Acetone	2.67	43	13980	10.289 ug/L	95
14) Carbon disulfide	2.81	76	22002	1.093 ug/L	99
21) 2-Butanone	4.74	43	3069	1.300 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082420\
Data File : VU039949.D
Acq On : 24 Aug 2020 16:11
Operator : SY/MD
Sample : L3767-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSG5

Quant Time: Aug 25 02:17:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 25 02:02:41 2020
Response via : Initial Calibration

