

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082421\
 Data File : VU044521.D
 Acq On : 24 Aug 2021 14:06
 Operator : SY/MD
 Sample : M3517-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AC9

Quant Time: Aug 25 03:51:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 25 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	114843	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	115321	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	50506	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	37011	3.985	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.600%
7) Chloroethane-d5	1.916	69	36505	4.522	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.400%
11) 1,1-Dichloroethene-d2	2.572	65	17363	4.394	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	4.639	46	131693	51.267	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.540%
24) Chloroform-d	5.070	84	77861	4.741	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	5.710	65	44245	4.940	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
32) Benzene-d6	5.736	84	156849	4.830	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.600%
36) 1,2-Dichloropropane-d6	6.697	67	51803	4.986	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.800%
41) Toluene-d8	7.903	98	129884	4.511	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
43) trans-1,3-Dichloroprop...	8.186	79	15258	4.090	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.800%
46) 2-Hexanone-d5	8.642	63	104471	51.781	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.560%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	42597	4.869	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	12.198	152	44124	5.171	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082421\
Data File : VU044521.D
Acq On : 24 Aug 2021 14:06
Operator : SY/MD
Sample : M3517-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AC9

Quant Time: Aug 25 03:51:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 25 03:50:31 2021
Response via : Initial Calibration

