

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082518\
 Data File : VU026383.D
 Acq On : 25 Aug 2018 19:54
 Operator : MD/SY
 Sample : VIBLK41
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK41

Quant Time: Aug 27 04:00:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 25 04:25:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	322013	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	319509	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	151884	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	100528	51.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.36%
7) Chloroethane-d5	1.68	69	80876	50.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	2.27	63	146048	37.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.66%
21) 2-Butanone-d5	4.18	46	146469	93.44	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.44%
24) Chloroform-d	4.65	84	200200	45.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.46%
26) 1,2-Dichloroethane-d4	5.31	65	136338	47.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.72%
32) Benzene-d6	5.35	84	398098	46.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.14%
36) 1,2-Dichloropropane-d6	6.33	67	128707	46.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.80%
41) Toluene-d8	7.57	98	362818	44.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.82%
43) trans-1,3-Dichloropropene-	7.85	79	59472	41.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.88%
47) 2-Hexanone-d5	8.31	63	110232	93.98	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.98%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	179172	44.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.22%
64) 1,2-Dichlorobenzene-d4	11.86	152	154009	47.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082518\
Data File : VU026383.D
Acq On : 25 Aug 2018 19:54
Operator : MD/SY
Sample : VIBLK41
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK41

Quant Time: Aug 27 04:00:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 25 04:25:01 2018
Response via : Initial Calibration

