

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082522\
 Data File : VU050442.D
 Acq On : 25 Aug 2022 12:41
 Operator : SY/MD
 Sample : N3734-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WS-OXYGENATES

Quant Time: Aug 26 15:17:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Aug 25 14:01:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.111	96	38268	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	9081	0.697	ug/l	0.00
Spiked Amount 1.000			Recovery	=	70.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	12587	0.817	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.581	101	164264	14.287	ug/l	97
15) Methylene Chloride	3.044	84	12576	0.825	ug/l	93
24) tert-Butyl Alcohol	3.202	59	51917	42.370	ug/l	# 1
25) Methyl tert-Butyl Ether	3.372	73	1176252	39.289	ug/l	94
31) Isopropyl Ether	4.005	45	1846799	49.061	ug/l	99
32) Ethyl-t-butyl ether	4.517	59	552474	18.682	ug/l	99
33) Tert-Amyl methyl ether	5.957	73	790400	41.000	ug/l	99
73) n-propylbenzene	10.909	120	184441	12.069	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082522\
Data File : VU050442.D
Acq On : 25 Aug 2022 12:41
Operator : SY/MD
Sample : N3734-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WS-OXYGENATES

Quant Time: Aug 26 15:17:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Aug 25 14:01:59 2022
Response via : Initial Calibration

