

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082522\
 Data File : VU050438.D
 Acq On : 25 Aug 2022 09:33
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 26 06:32:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Aug 25 14:01:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.414	0.357	13.8	90	0.00
3 t	Chloromethane	0.504	0.452	10.3	96	0.00
4 Rt	Vinyl Chloride	0.431	0.389	9.7	93	0.00
5 T	Bromomethane	0.176	0.162	8.0	102	0.00
6 T	Chloroethane	0.260	0.238	8.5	95	0.00
7 T	Trichlorofluoromethane	0.537	0.472	12.1	91	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.300	0.269	10.3	95	0.00
9 Rt	1,1-Dichloroethene	0.295	0.265	10.2	94	0.00
10 t	Iodomethane	0.243	0.262	-7.8	109	0.00
11 t	Allyl Chloride	0.454	0.433	4.6	99	0.00
12 t	Acrylonitrile	0.088	0.093	-5.7	108	0.00
13 T	Acetone	0.062	0.056	9.7	96	0.00
14 T	Carbon Disulfide	0.912	0.766	16.0	89	0.00
15 RT	Methylene Chloride	0.398	0.351	11.8	104	0.00
16 RT	trans-1,2-Dichloroethene	0.304	0.290	4.6	99	0.00
17 t	1,1-Dichloroethane	0.634	0.610	3.8	101	0.00
18 T	2-Butanone	0.097	0.098	-1.0	101	0.00
19	Cyclohexane	0.460	0.378	17.8	87	0.00
20	Methylcyclohexane	0.363	0.405	-11.6	101	0.00
21 T	2,2-Dichloropropane	0.482	0.411	14.7	95	0.00
22 RT	cis-1,2-Dichloroethene	0.344	0.339	1.5	103	0.00
23 t	Diethyl Ether	0.253	0.253	0.0	104	0.00
24 t	tert-Butyl Alcohol	0.032	0.031	3.1	109	0.00
25 t	Methyl tert-Butyl Ether	0.782	0.793	-1.4	107	0.00
26 t	Bromochloromethane	0.138	0.160	-15.9	125	0.00
27 t	Chloroform	0.614	0.599	2.4	103	0.00
28 RT	1,1,1-Trichloroethane	0.524	0.460	12.2	92	0.00
29 T	1,1-Dichloropropene	0.420	0.392	6.7	99	0.00
30 RT	Carbon Tetrachloride	0.420	0.379	9.8	92	0.00
31 t	Isopropyl Ether	0.984	0.984	0.0	106	0.00
32	Ethyl-t-butyl ether	0.773	0.684	11.5	90	0.00
33	Tert-Amyl methyl ether	0.504	0.577	-14.5	101	0.00
34 t	Propionitrile	0.029	0.032	-10.3	133	0.00
35 RT	Benzene	1.234	1.267	-2.7	104	0.00
36 RT	1,2-Dichloroethane	0.377	0.373	1.1	103	0.00
37 RT	Trichloroethene	0.296	0.299	-1.0	105	0.00
38 Rt	1,2-Dichloropropane	0.331	0.351	-6.0	105	0.00
39 t	Methacrylonitrile	0.118	0.106	10.2	103	0.00
40 t	Methyl acrylate	0.179	0.184	-2.8	119	0.00
41 t	Tetrahydrofuran	0.059	0.058	1.7	103	0.00
42 t	1-Chlorobutane	0.598	0.546	8.7	98	0.00
43 T	Dibromomethane	0.179	0.185	-3.4	105	0.00
44 T	Bromodichloromethane	0.421	0.438	-4.0	106	0.00
45 T	4-Methyl-2-Pentanone	0.179	0.219	-22.3	110	0.00
46 t	t-1,4-Dichloro-2-butene	0.090	0.093	-3.3	88	0.00
47 t	Methyl methacrylate	0.134	0.170	-26.9	109	0.00
48 t	Ethyl methacrylate	0.244	0.303	-24.2	108	0.00
49 Rt	Toluene	0.663	0.772	-16.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082522\
 Data File : VU050438.D
 Acq On : 25 Aug 2022 09:33
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 26 06:32:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Aug 25 14:01:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.328	0.383	-16.8	110	0.00
51 T	cis-1,3-Dichloropropene	0.376	0.442	-17.6	112	0.00
52 RT	1,1,2-Trichloroethane	0.263	0.271	-3.0	104	0.00
53 t	1,3-Dichloropropane	0.413	0.455	-10.2	107	0.00
54 t	2-Hexanone	0.125	0.154	-23.2	112	0.00
55 t	Dibromochloromethane	0.289	0.303	-4.8	107	0.00
56 T	1,2-Dibromoethane	0.236	0.247	-4.7	104	0.00
57 S	4-Bromofluorobenzene	0.340	0.473	-39.1#	129	0.00
58 RT	Tetrachloroethene	0.278	0.267	4.0	98	0.00
59 Rt	Chlorobenzene	0.716	0.808	-12.8	106	0.00
60 T	1,1,1,2-Tetrachloroethane	0.301	0.303	-0.7	101	0.00
61 t	Pentachloroethane	0.250	0.255	-2.0	102	0.00
62 t	Hexachloroethane	0.208	0.226	-8.7	104	0.00
63 Rt	Ethyl Benzene	1.140	1.395	-22.4	104	0.00
64 RT	m/p-Xylenes	0.444	0.555	-25.0	101	0.00
65 RT	o-Xylene	0.423	0.527	-24.6	104	0.00
66 RT	Styrene	0.720	0.928	-28.9	103	0.00
67 t	Bromoform	0.164	0.177	-7.9	108	0.00
68 S	1,2-Dichlorobenzene-d4	0.402	0.408	-1.5	103	0.00
69 T	Isopropylbenzene	1.100	1.361	-23.7	104	0.00
70 T	1,1,2,2-Tetrachloroethane	0.361	0.381	-5.5	107	0.00
71 T	1,2,3-Trichloropropane	0.272	0.261	4.0	94	0.00
72 t	Bromobenzene	0.309	0.343	-11.0	104	0.00
73 t	n-propylbenzene	0.300	0.361	-20.3	97	0.00
74 t	2-Chlorotoluene	0.291	0.340	-16.8	100	0.00
75 t	1,3,5-Trimethylbenzene	0.973	1.236	-27.0	102	0.00
76 t	4-Chlorotoluene	0.298	0.340	-14.1	93	0.00
77 t	tert-Butylbenzene	0.957	1.146	-19.7	102	0.00
78 t	1,2,4-Trimethylbenzene	1.003	1.267	-26.3	102	0.00
79 t	sec-Butylbenzene	1.258	1.505	-19.6	99	0.00
80	Nitrobenzene	0.020	0.020	0.0	106	0.00
81 t	p-Isopropyltoluene	1.003	1.232	-22.8	98	0.00
82 t	1,3-Dichlorobenzene	0.636	0.674	-6.0	98	0.00
83 Rt	1,4-Dichlorobenzene	0.633	0.695	-9.8	101	0.00
84 t	n-Butylbenzene	0.948	1.040	-9.7	93	0.00
85 Rt	1,2-Dichlorobenzene	0.625	0.681	-9.0	102	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.057	0.061	-7.0	108	0.00
87 Rt	1,2,4-Trichlorobenzene	0.357	0.381	-6.7	96	0.00
88 t	Hexachlorobutadiene	0.232	0.225	3.0	95	0.00
89 t	Naphthalene	0.786	0.842	-7.1	99	0.00
90 t	1,2,3-Trichlorobenzene	0.376	0.407	-8.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0