

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082522\
 Data File : VU050447.D
 Acq On : 25 Aug 2022 18:22
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 26 06:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Aug 25 14:01:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.414	0.387	6.5	90	0.00
3 t	Chloromethane	0.504	0.467	7.3	91	0.00
4 Rt	Vinyl Chloride	0.431	0.417	3.2	92	0.00
5 T	Bromomethane	0.176	0.163	7.4	94	0.00
6 T	Chloroethane	0.260	0.261	-0.4	96	0.00
7 T	Trichlorofluoromethane	0.537	0.527	1.9	94	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.300	0.293	2.3	95	0.00
9 Rt	1,1-Dichloroethene	0.295	0.288	2.4	94	0.00
10 t	Iodomethane	0.243	0.248	-2.1	95	0.00
11 t	Allyl Chloride	0.454	0.462	-1.8	98	0.00
12 t	Acrylonitrile	0.088	0.095	-8.0	102	0.00
13 T	Acetone	0.062	0.066	-6.5	104	0.00
14 T	Carbon Disulfide	0.912	0.810	11.2	87	0.00
15 RT	Methylene Chloride	0.398	0.383	3.8	105	0.00
16 RT	trans-1,2-Dichloroethene	0.304	0.317	-4.3	100	0.00
17 t	1,1-Dichloroethane	0.634	0.612	3.5	94	0.00
18 T	2-Butanone	0.097	0.094	3.1	90	0.00
19	Cyclohexane	0.460	0.398	13.5	85	0.00
20	Methylcyclohexane	0.363	0.417	-14.9	96	0.00
21 T	2,2-Dichloropropane	0.482	0.416	13.7	89	0.00
22 RT	cis-1,2-Dichloroethene	0.344	0.362	-5.2	102	0.00
23 t	Diethyl Ether	0.253	0.269	-6.3	102	0.00
24 t	tert-Butyl Alcohol	0.032	0.037	-15.6	121	0.00
25 t	Methyl tert-Butyl Ether	0.782	0.846	-8.2	105	0.00
26 t	Bromochloromethane	0.138	0.167	-21.0	120	0.00
27 t	Chloroform	0.614	0.653	-6.4	103	0.00
28 RT	1,1,1-Trichloroethane	0.524	0.502	4.2	93	0.00
29 T	1,1-Dichloropropene	0.420	0.418	0.5	98	0.00
30 RT	Carbon Tetrachloride	0.420	0.410	2.4	92	0.00
31 t	Isopropyl Ether	0.984	0.825	16.2	82	0.00
32	Ethyl-t-butyl ether	0.773	0.677	12.4	82	0.00
33	Tert-Amyl methyl ether	0.504	0.608	-20.6	99	0.00
34 t	Propionitrile	0.029	0.032	-10.3	123	0.00
35 RT	Benzene	1.234	1.338	-8.4	101	0.00
36 RT	1,2-Dichloroethane	0.377	0.416	-10.3	106	0.00
37 RT	Trichloroethene	0.296	0.317	-7.1	103	0.00
38 Rt	1,2-Dichloropropane	0.331	0.384	-16.0	106	0.00
39 t	Methacrylonitrile	0.118	0.105	11.0	94	0.00
40 t	Methyl acrylate	0.179	0.163	8.9	98	0.00
41 t	Tetrahydrofuran	0.059	0.056	5.1	92	0.00
42 t	1-Chlorobutane	0.598	0.579	3.2	96	0.00
43 T	Dibromomethane	0.179	0.197	-10.1	103	0.00
44 T	Bromodichloromethane	0.421	0.480	-14.0	107	0.00
45 T	4-Methyl-2-Pentanone	0.179	0.222	-24.0	103	0.00
46 t	t-1,4-Dichloro-2-butene	0.090	0.088	2.2	76	0.00
47 t	Methyl methacrylate	0.134	0.175	-30.6#	104	0.00
48 t	Ethyl methacrylate	0.244	0.307	-25.8	101	0.00
49 Rt	Toluene	0.663	0.826	-24.6	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082522\
 Data File : VU050447.D
 Acq On : 25 Aug 2022 18:22
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Aug 26 06:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Aug 25 14:01:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.328	0.400	-22.0	106	0.00
51 T	cis-1,3-Dichloropropene	0.376	0.454	-20.7	106	0.00
52 RT	1,1,2-Trichloroethane	0.263	0.296	-12.5	105	0.00
53 t	1,3-Dichloropropane	0.413	0.489	-18.4	106	0.00
54 t	2-Hexanone	0.125	0.153	-22.4	103	0.00
55 t	Dibromochloromethane	0.289	0.333	-15.2	109	0.00
56 T	1,2-Dibromoethane	0.236	0.264	-11.9	102	0.00
57 S	4-Bromofluorobenzene	0.340	0.387	-13.8	97	0.00
58 RT	Tetrachloroethene	0.278	0.288	-3.6	98	0.00
59 Rt	Chlorobenzene	0.716	0.860	-20.1	104	0.00
60 T	1,1,1,2-Tetrachloroethane	0.301	0.331	-10.0	102	0.00
61 t	Pentachloroethane	0.250	0.273	-9.2	101	0.00
62 t	Hexachloroethane	0.208	0.248	-19.2	105	0.00
63 Rt	Ethyl Benzene	1.140	1.468	-28.8	101	0.00
64 RT	m/p-Xylenes	0.444	0.601	-35.4#	101	0.00
65 RT	o-Xylene	0.423	0.560	-32.4#	102	0.00
66 RT	Styrene	0.720	0.999	-38.8#	102	0.00
67 t	Bromoform	0.164	0.188	-14.6	106	0.00
68 S	1,2-Dichlorobenzene-d4	0.402	0.428	-6.5	100	0.00
69 T	Isopropylbenzene	1.100	1.412	-28.4	99	0.00
70 T	1,1,2,2-Tetrachloroethane	0.361	0.407	-12.7	106	0.00
71 T	1,2,3-Trichloropropane	0.272	0.262	3.7	87	0.00
72 t	Bromobenzene	0.309	0.367	-18.8	102	0.00
73 t	n-propylbenzene	0.300	0.396	-32.0#	98	0.00
74 t	2-Chlorotoluene	0.291	0.374	-28.5	101	0.00
75 t	1,3,5-Trimethylbenzene	0.973	1.295	-33.1#	98	0.00
76 t	4-Chlorotoluene	0.298	0.365	-22.5	92	0.00
77 t	tert-Butylbenzene	0.957	1.191	-24.5	98	0.00
78 t	1,2,4-Trimethylbenzene	1.003	1.372	-36.8#	102	0.00
79 t	sec-Butylbenzene	1.258	1.640	-30.4#	99	0.00
80	Nitrobenzene	0.020	0.017	15.0	83	0.00
81 t	p-Isopropyltoluene	1.003	1.306	-30.2#	96	0.00
82 t	1,3-Dichlorobenzene	0.636	0.757	-19.0	101	0.00
83 Rt	1,4-Dichlorobenzene	0.633	0.767	-21.2	102	0.00
84 t	n-Butylbenzene	0.948	1.074	-13.3	88	0.00
85 Rt	1,2-Dichlorobenzene	0.625	0.751	-20.2	103	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.057	0.061	-7.0	100	0.00
87 Rt	1,2,4-Trichlorobenzene	0.357	0.387	-8.4	90	0.00
88 t	Hexachlorobutadiene	0.232	0.236	-1.7	91	0.00
89 t	Naphthalene	0.786	0.751	4.5	82	0.00
90 t	1,2,3-Trichlorobenzene	0.376	0.409	-8.8	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0