

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082622\
 Data File : VU050451.D
 Acq On : 26 Aug 2022 12:51
 Operator : SY/MD
 Sample : VU0826WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK168

Quant Time: Aug 27 02:45:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM082422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 27 02:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	302336	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	305224	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	118573	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	164065	45.893	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	91.780%	
7) Chloroethane-d5	1.887	69	125577	47.473	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery	=	94.940%	
11) 1,1-Dichloroethene-d2	2.555	63	189904	37.589	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	75.180%	
21) 2-Butanone-d5	4.620	46	234547	100.237	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	100.240%	
24) Chloroform-d	5.060	84	260812	45.448	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.900%	
26) 1,2-Dichloroethane-d4	5.700	65	169123	47.419	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.840%	
32) Benzene-d6	5.726	84	542649	48.116	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.240%	
36) 1,2-Dichloropropane-d6	6.690	67	177146	47.414	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	94.820%	
41) Toluene-d8	7.896	98	443058	44.897	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.800%	
43) trans-1,3-Dichloroprop...	8.179	79	67814	43.513	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.020%	
47) 2-Hexanone-d5	8.632	63	139976	100.957	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	100.960%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	275724	47.847	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	95.700%	
66) 1,2-Dichlorobenzene-d4	12.195	152	152984	54.011	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	108.020%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082622\
Data File : VU050451.D
Acq On : 26 Aug 2022 12:51
Operator : SY/MD
Sample : VU0826WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK168

Quant Time: Aug 27 02:45:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM082422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 27 02:44:18 2022
Response via : Initial Calibration

