

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082622\
 Data File : VU050452.D
 Acq On : 26 Aug 2022 13:15
 Operator : SY/MD
 Sample : N4362-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0FB6

Quant Time: Aug 27 02:45:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM082422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 27 02:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	280079	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	287771	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	105051	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	148680	44.895	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.780%	
7) Chloroethane-d5	1.899	69	127654	52.093	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	104.180%	
11) 1,1-Dichloroethene-d2	2.562	63	173561	37.084	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.160%	
21) 2-Butanone-d5	4.623	46	229727	105.979	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	105.980%	
24) Chloroform-d	5.063	84	262817	49.437	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.880%	
26) 1,2-Dichloroethane-d4	5.700	65	172225	52.126	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.260%	
32) Benzene-d6	5.726	84	531020	49.941	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.880%	
36) 1,2-Dichloropropane-d6	6.690	67	178703	50.732	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.460%	
41) Toluene-d8	7.899	98	427981	45.999	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	92.000%	
43) trans-1,3-Dichloroprop...	8.179	79	67167	45.712	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	91.420%	
47) 2-Hexanone-d5	8.632	63	126497	96.768	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	96.770%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	264340	48.654	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.300%	
66) 1,2-Dichlorobenzene-d4	12.192	152	138913	55.356	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.720%	
Target Compounds						Qvalue
3) Chloromethane	1.520	50	3732	1.465	ug/L	90
13) Acetone	2.626	43	8798	6.173	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082622\
Data File : VU050452.D
Acq On : 26 Aug 2022 13:15
Operator : SY/MD
Sample : N4362-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0FB6

Quant Time: Aug 27 02:45:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM082422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 27 02:44:18 2022
Response via : Initial Calibration

