

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082718\
 Data File : VU026394.D
 Acq On : 27 Aug 2018 13:14
 Operator : MD/SY
 Sample : VIBLK46
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK46

Quant Time: Aug 27 14:59:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Aug 27 13:46:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	335306	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	326915	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	153637	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	99505	48.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.30%
7) Chloroethane-d5	1.68	69	81556	48.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.27	63	148301	36.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.78%
21) 2-Butanone-d5	4.18	46	151312	92.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.70%
24) Chloroform-d	4.65	84	210598	46.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.31	65	139877	47.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.30%
32) Benzene-d6	5.35	84	409521	46.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.64%
36) 1,2-Dichloropropane-d6	6.33	67	131290	46.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.50%
41) Toluene-d8	7.57	98	371768	44.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.94%
43) trans-1,3-Dichloropropene-	7.85	79	55014	37.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.84%
47) 2-Hexanone-d5	8.31	63	115336	96.10	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.10%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	179809	43.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	159012	48.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082718\
Data File : VU026394.D
Acq On : 27 Aug 2018 13:14
Operator : MD/SY
Sample : VIBLK46
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK46

Quant Time: Aug 27 14:59:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 27 13:46:46 2018
Response via : Initial Calibration

