

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082719\
 Data File : VU033964.D
 Acq On : 26 Aug 2019 22:11
 Operator : JC/SP
 Sample : K4539-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP

Quant Time: Aug 27 12:21:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 27 06:58:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	416060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	671758	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	624604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	266257	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.28	65	299712	50.13	ug/l	0.00
Spiked Amount						
			Recovery	=	100.26%	
35) Dibromofluoromethane	4.86	113	225885	48.34	ug/l	0.00
Spiked Amount						
			Recovery	=	96.68%	
50) Toluene-d8	7.55	98	875696	50.66	ug/l	0.00
Spiked Amount						
			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	10.29	95	276656	41.52	ug/l	0.00
Spiked Amount						
			Recovery	=	83.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082719\
Data File : VU033964.D
Acq On : 26 Aug 2019 22:11
Operator : JC/SP
Sample : K4539-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP

Quant Time: Aug 27 12:21:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082719W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 27 06:58:46 2019
Response via : Initial Calibration

