

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082719\
 Data File : VU033967.D
 Acq On : 26 Aug 2019 23:24
 Operator : JC/SP
 Sample : K4515-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB-1-082319

Quant Time: Sep 05 08:45:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 27 06:58:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	402567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	655226	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	638480	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	267804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.28	65	294984	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
35) Dibromofluoromethane	4.86	113	224170	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
50) Toluene-d8	7.54	98	822328	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
62) 4-Bromofluorobenzene	10.29	95	276796	42.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.18%	
Target Compounds						
16) Acetone	2.31	43	13736	3.989	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082719\
Data File : VU033967.D
Acq On : 26 Aug 2019 23:24
Operator : JC/SP
Sample : K4515-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB-1-082319

Quant Time: Sep 05 08:45:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082719W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 27 06:58:46 2019
Response via : Initial Calibration

