

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082719\
 Data File : VU033968.D
 Acq On : 26 Aug 2019 23:48
 Operator : JC/SP
 Sample : K4539-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PZ-1

Quant Time: Aug 27 12:23:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 27 06:58:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.95	168	401329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	655163	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	632135	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	262031	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.28	65	296659	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	4.86	113	225397	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	7.55	98	851532	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
62) 4-Bromofluorobenzene	10.29	95	260429	40.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.16%	
Target Compounds						
16) Acetone	2.31	43	11060	3.222	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082719\
Data File : VU033968.D
Acq On : 26 Aug 2019 23:48
Operator : JC/SP
Sample : K4539-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PZ-1

Quant Time: Aug 27 12:23:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082719W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 27 06:58:46 2019
Response via : Initial Calibration

