

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082720\
 Data File : VU039987.D
 Acq On : 27 Aug 2020 12:06
 Operator : SY/MD
 Sample : VU0827WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK27

Quant Time: Aug 28 01:18:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 28 01:12:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	71312	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	69915	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	32432	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	24760	3.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.60%
7) Chloroethane-d5	1.92	69	22918	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.58	63	40108	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	4.66	46	95052	47.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.50%
24) Chloroform-d	5.08	84	55475	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	5.72	65	34585	5.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.00%
32) Benzene-d6	5.74	84	98512	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.70	67	33598	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.91	98	84506	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
43) trans-1,3-Dichloropropene-	8.19	79	14023	4.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.40%
46) 2-Hexanone-d5	8.64	63	68500	46.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.70%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	27961	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	35603	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1710	0.199 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082720\
Data File : VU039987.D
Acq On : 27 Aug 2020 12:06
Operator : SY/MD
Sample : VU0827WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK27

Quant Time: Aug 28 01:18:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 28 01:12:06 2020
Response via : Initial Calibration

