

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082721\
 Data File : VU044528.D
 Acq On : 27 Aug 2021 11:33
 Operator : SY/MD
 Sample : M3554-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AH9

Quant Time: Aug 28 01:20:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 28 01:19:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	128810	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	129587	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	59292	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	30249	2.904	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	58.000%
7) Chloroethane-d5	1.919	69	31716	3.503	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.000%
11) 1,1-Dichloroethene-d2	2.572	65	14835	3.348	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.000%
20) 2-Butanone-d5	4.639	46	123678	42.927	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.860%
24) Chloroform-d	5.073	84	72199	3.919	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.400%
26) 1,2-Dichloroethane-d4	5.710	65	40730	4.055	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.000%
32) Benzene-d6	5.735	84	141547	3.879	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
36) 1,2-Dichloropropane-d6	6.700	67	46790	4.008	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.200%
41) Toluene-d8	7.906	98	118086	3.649	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.000%
43) trans-1,3-Dichloroprop...	8.185	79	15395	3.672	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.400%
46) 2-Hexanone-d5	8.645	63	97294	42.915	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.820%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	38841	3.951	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	12.198	152	42078	4.200	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082721\
Data File : VU044528.D
Acq On : 27 Aug 2021 11:33
Operator : SY/MD
Sample : M3554-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH9

Quant Time: Aug 28 01:20:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 28 01:19:29 2021
Response via : Initial Calibration

