

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082721\
 Data File : VU044530.D
 Acq On : 27 Aug 2021 12:22
 Operator : SY/MD
 Sample : M3554-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AJ1

Quant Time: Aug 28 01:20:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 28 01:19:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	113379	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	115973	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	51265	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	32490	3.543	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.800%
7) Chloroethane-d5	1.916	69	32877	4.125	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.600%
11) 1,1-Dichloroethene-d2	2.571	65	15878	4.071	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.400%
20) 2-Butanone-d5	4.639	46	129765	51.169	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.340%
24) Chloroform-d	5.073	84	75312	4.645	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.710	65	43161	4.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	5.735	84	149775	4.586	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
36) 1,2-Dichloropropane-d6	6.697	67	49765	4.763	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.906	98	125865	4.346	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.000%
43) trans-1,3-Dichloroprop...	8.185	79	16278	4.338	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.800%
46) 2-Hexanone-d5	8.642	63	102297	50.418	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.840%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	42120	4.787	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.198	152	43521	5.025	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082721\
Data File : VU044530.D
Acq On : 27 Aug 2021 12:22
Operator : SY/MD
Sample : M3554-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AJ1

Quant Time: Aug 28 01:20:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 28 01:19:29 2021
Response via : Initial Calibration

