

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082819\
 Data File : VU034001.D
 Acq On : 28 Aug 2019 03:51
 Operator : JC/SP
 Sample : K4516-33
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T2-S-B3-(0)

Quant Time: Aug 28 08:44:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 28 02:42:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	449179	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	721556	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	694349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	287622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	304021	60.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.84%	
35) Dibromofluoromethane	4.86	113	239215	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	7.55	98	929840	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	10.29	95	310048	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
16) Acetone	2.31	43	12003	5.025	ug/l	97
20) Methylene Chloride	2.69	84	17689	3.491	ug/l	92
43) Isopropyl Acetate	5.53	43	90994	8.953	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082819\
Data File : VU034001.D
Acq On : 28 Aug 2019 03:51
Operator : JC/SP
Sample : K4516-33
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T2-S-B3-(0)

Quant Time: Aug 28 08:44:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082819W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 28 02:42:03 2019
Response via : Initial Calibration

