

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082918\
 Data File : VU026430.D
 Acq On : 29 Aug 2018 18:37
 Operator : MD/SY
 Sample : J4691-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEDZ9

Quant Time: Aug 30 02:45:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 30 02:20:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	150757	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	150195	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	71521	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	43493	46.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.12%
7) Chloroethane-d5	1.68	69	41083	48.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.40%
11) 1,1-Dichloroethene-d2	2.27	63	72168	34.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.62%
21) 2-Butanone-d5	4.18	46	71630	98.81	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.81%
24) Chloroform-d	4.65	84	83205	47.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.32%
26) 1,2-Dichloroethane-d4	5.31	65	58212	49.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.12%
32) Benzene-d6	5.35	84	168256	48.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.33	67	57045	47.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.94%
41) Toluene-d8	7.57	98	159781	47.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.62%
43) trans-1,3-Dichloropropene-	7.85	79	26164	48.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.36%
47) 2-Hexanone-d5	8.31	63	55820	102.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.34%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	78708	44.92	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.84%
64) 1,2-Dichlorobenzene-d4	11.86	152	67321	51.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.76%

Target Compounds						Ovalue
13) Acetone	2.32	43	4405	7.02	ug/L	98
34) Trichloroethene	6.19	95	2039	1.89	ug/L	90
46) Tetrachloroethene	8.24	164	2428240	2749.89	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082918\
Data File : VU026430.D
Acq On : 29 Aug 2018 18:37
Operator : MD/SY
Sample : J4691-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEDZ9

Quant Time: Aug 30 02:45:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 30 02:20:47 2018
Response via : Initial Calibration

