

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU082919\
 Data File : VU034028.D
 Acq On : 28 Aug 2019 17:19
 Operator : JC/SP
 Sample : K4592-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5

Quant Time: Aug 29 04:17:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	477327	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	756278	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	708578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	345777	50.00	ug/l	0.00


System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.29	65	291447	47.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.30%	
35) Dibromofluoromethane	4.86	113	243792	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	
50) Toluene-d8	7.55	98	940063	47.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.72%	
62) 4-Bromofluorobenzene	10.29	95	329094	43.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.14%	

Target Compounds

					Qvalue	
3) Chloromethane	1.32	50	2588	Below Cal		89
5) Bromomethane	1.58	94	112	Below Cal	#	2
11) Tert butyl alcohol	2.82	59	6084	4.314	ug/l #	85
13) Acrolein	2.19	56	425	0.716	ug/l	87
16) Acetone	2.31	43	10946	3.507	ug/l	93
17) Carbon Disulfide	2.47	76	11785	0.779	ug/l	96
18) Methyl Acetate	2.43	43	22253	2.759	ug/l #	51
20) Methylene Chloride	2.68	84	1493	0.253	ug/l	95
31) Cyclohexane	4.96	56	8684	0.951	ug/l #	1
36) 1,1-Dichloropropene	4.96	75	34408	4.810	ug/l #	48
38) Carbon Tetrachloride	4.96	117	45559	6.598	ug/l #	17
49) 1,4-Dioxane	6.61	88	409	2.776	ug/l #	72
51) 4-Methyl-2-Pentanone	7.45	43	4067	0.483	ug/l	98
56) Ethyl methacrylate	8.03	69	491	1.023	ug/l #	77
59) 2-Hexanone	8.38	43	4155	0.654	ug/l	76
74) N-amyl acetate	10.01	43	2148	0.220	ug/l #	68
76) 1,2,3-Trichloropropane	10.29	75	157753	20.242	ug/l #	37
81) trans-1,4-Dichloro-2-buten	10.29	75	157753	68.161	ug/l #	100
82) 4-Chlorotoluene	10.76	91	3864	0.207	ug/l	93
87) 1,3-Dichlorobenzene	11.40	146	3107	0.281	ug/l	96
88) 1,4-Dichlorobenzene	11.40	146	3107	0.278	ug/l	96
89) n-Butylbenzene	11.88	91	4918	0.266	ug/l	86
91) 1,2-Dichlorobenzene	11.87	146	2766	0.261	ug/l	89
93) 1,2,4-Trichlorobenzene	13.50	180	2026	1.377	ug/l	89
94) Hexachlorobutadiene	13.68	225	1478	0.357	ug/l	94
95) Naphthalene	13.75	128	7751	2.136	ug/l #	95
96) 1,2,3-Trichlorobenzene	13.98	180	4048	0.596	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. 0.00 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampleId :


GW-5

Tgt Ion:168 Resp: 477327

Ion Ratio Lower Upper

168 100

99 54.1 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): VU034018.D (-1185) (-)

Ion 98.90 (98.60 to 99.60): VU034018.D (-1185) (-)

4.96

250000

200000

150000



100000

50000

0

4.90 4.95 5.00 5.05

Time-->



#3

Chloromethane

Concen: Below Cal

RT: 1.32 min Scan# 73

Delta R.T. 0.00 min

Lab File: VU034028.D


Acq: 28 Aug 2019 17:19

Tgt Ion: 50 Resp: 2588

Ion Ratio Lower Upper

50 100

52 39.1 26.2 39.4



Abundance Ion 49.90 (49.60 to 50.60): VU034018.D (-63) (-)

Ion 51.90 (51.60 to 52.60): VU034018.D (-63) (-)

1.32

3000


2000


1000

0

1.30 1.32 1.34 1.36

Time-->





#5

Bromomethane

Concen: Below Cal

RT: 1.58 min Scan# 152

Delta R.T. -0.02 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampleId :





GW-5

Tgt Ion: 94 Resp: 112

Ion Ratio Lower Upper

94 100

96 0.0 76.3 114.5#



#11

Tert butyl alcohol

Concen: 4.314 ug/l

RT: 2.82 min Scan# 537

Delta R.T. -0.00 min

Lab File: VU034028.D




Acq: 28 Aug 2019 17:19


Tgt Ion: 59 Resp: 6084

Ion Ratio Lower Upper

59 100

57 4.5 8.0 12.0#





#13

Acrolein

Concen: 0.716 ug/l

RT: 2.19 min Scan# 341

Delta R.T. 0.01 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampleId :


GW-5

Tgt Ion: 56 Resp: 425




Ion Ratio Lower Upper

56 100

55 82.6 57.1 85.7



Abundance Ion 56.00 (55.70 to 56.70): VU034018.D (-327) (-)



#16

Acetone

Concen: 3.507 ug/l

RT: 2.31 min Scan# 380

Delta R.T. 0.00 min

Lab File: VU034028.D


Acq: 28 Aug 2019 17:19

Tgt Ion: 43 Resp: 10946



Ion Ratio Lower Upper


43 100

58 29.8 26.8 40.2



Abundance Ion 43.00 (42.70 to 43.70): VU034018.D (-358) (-)





#17

Carbon Disulfide

Concen: 0.779 ug/l

RT: 2.47 min Scan# 430

Delta R.T. 0.01 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampleId :


GW-5

Tgt Ion: 76 Resp: 11785

Ion Ratio Lower Upper

76 100

78 7.6 7.1 10.7



Abundance Ion 75.90 (75.60 to 76.60): VU034028.D (-412) (-)

Ion 77.90 (77.60 to 78.60): VU034028.D (-412) (-)

2.47

6000

5000

4000



3000

2000

1000

0

Time-->



#18

Methyl Acetate

Concen: 2.759 ug/l

RT: 2.43 min Scan# 417

Delta R.T. -0.17 min

Lab File: VU034028.D


Acq: 28 Aug 2019 17:19

Tgt Ion: 43 Resp: 22253

Ion Ratio Lower Upper

43 100

74 0.0 19.2 28.8#



Abundance Ion 43.00 (42.70 to 43.70): VU034028.D (-376) (-)

Ion 74.00 (73.70 to 74.70): VU034028.D (-376) (-)


2.43


10000

5000

0

Time-->









#20
Methylene Chloride
Concen: 0.253 ug/l
RT: 2.68 min Scan# 496
Delta R.T. 0.00 min
Lab File: VU034028.D
Acq: 28 Aug 2019 17:19




Instrument :
MSVOA_U
ClientSampleId :
GW-5


Tgt Ion:	84	Resp:	1493
Ion	Ratio	Lower	Upper
84	100		
49	132.8	100.2	150.4
51	36.7	30.8	46.2
86	63.6	52.5	78.7



#31
Cyclohexane
Concen: 0.951 ug/l
RT: 4.96 min Scan# 1204
Delta R.T. -0.01 min
Lab File: VU034028.D
Acq: 28 Aug 2019 17:19

Tgt Ion:	56	Resp:	8684
Ion	Ratio	Lower	Upper
56	100		
69	197.6	25.4	38.0#
84	188.4	65.8	98.8#









#33
 1,2-Dichloroethane-d4
 Concen: 47.652 ug/l
 RT: 5.29 min Scan# 1306
 Delta R.T. 0.00 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampled :
 GW-5


Tgt Ion: 65 Resp: 291447
 Ion Ratio Lower Upper
 65 100
 67 54.9 0.0 110.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.86 min Scan# 1485
 Delta R.T. 0.00 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion: 114 Resp: 756278
 Ion Ratio Lower Upper
 114 100
 63 21.4 0.0 42.0
 88 16.1 0.0 32.0









#35
 Dibromofluoromethane
 Concen: 46.477 ug/l
 RT: 4.86 min Scan# 1173
 Delta R.T. 0.00 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampled :
 GW-5


Tgt Ion:113 Resp: 243792
 Ion Ratio Lower Upper
 113 100
 111 101.7 81.7 122.5
 192 21.0 17.1 25.7



#36
 1,1-Dichloropropene
 Concen: 4.810 ug/l
 RT: 4.96 min Scan# 1204
 Delta R.T. -0.15 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion: 75 Resp: 34408
 Ion Ratio Lower Upper
 75 100
 110 8.1 17.8 53.5#
 77 0.0 25.0 37.4#





#38

Carbon Tetrachloride

Concen: 6.598 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. -0.15 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampled :

GW-5


Tgt Ion:117 Resp: 45559

Ion Ratio Lower Upper

117 100

119 6.0 77.0 115.4#




121 0.0 23.9 35.9#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#49

1,4-Dioxane

Concen: 2.776 ug/l

RT: 6.61 min Scan# 1718

Delta R.T. 0.02 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19


Tgt Ion: 88 Resp: 409

Ion Ratio Lower Upper

88 100

43 43.0 24.8 37.2#



58 48.2 59.3 88.9#




Abundance Ion 88.00 (87.70 to 88.70): VU0

Ion 43.00 (42.70 to 43.70): VU0

Ion 58.00 (57.70 to 58.70): VU0









#50
Toluene-d8
Concen: 47.862 ug/l
RT: 7.55 min Scan# 2008
Delta R.T. 0.00 min
Lab File: VU034028.D
Acq: 28 Aug 2019 17:19




Instrument :
MSVOA_U
ClientSampleId :
GW-5


Tgt Ion: 98 Resp: 940063
Ion Ratio Lower Upper
98 100
100 63.7 50.3 75.5



#51
4-Methyl-2-Pentanone
Concen: 0.483 ug/l
RT: 7.45 min Scan# 1979
Delta R.T. 0.01 min
Lab File: VU034028.D
Acq: 28 Aug 2019 17:19

Tgt Ion: 43 Resp: 4067
Ion Ratio Lower Upper
43 100
58 38.9 32.2 48.4









#56
 Ethyl methacrylate
 Concen: 1.023 ug/l
 RT: 8.03 min Scan# 2158
 Delta R.T. 0.03 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5


Tgt Ion: 69 Resp: 491
 Ion Ratio Lower Upper
 69 100
 41 77.4 50.7 76.1#
 39 12.4 24.2 36.2#



#59
 2-Hexanone
 Concen: 0.654 ug/l
 RT: 8.38 min Scan# 2267
 Delta R.T. 0.04 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion: 43 Resp: 4155
 Ion Ratio Lower Upper
 43 100
 58 39.0 28.4 85.3





#62

4-Bromofluorobenzene

Concen: 43.074 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. 0.00 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampled :

GW-5


Tgt Ion: 95 Resp: 329094

Ion Ratio Lower Upper

95 100

174 83.6 0.0 168.8




176 80.6 0.0 161.8



Abundance Ion 94.90 (94.60 to 95.60): VU0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.07 min Scan# 2482

Delta R.T. 0.00 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19


Tgt Ion: 117 Resp: 708578

Ion Ratio Lower Upper

117 100

82 56.6 45.1 67.7



119 31.6 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VU0

Ion 118.90 (118.60 to 119.60): V









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 11.47 min Scan# 3227
 Delta R.T. 0.00 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5


Tgt Ion:152 Resp: 345777
 Ion Ratio Lower Upper
 152 100
 115 57.1 41.1 123.3
 150 155.2 0.0 348.8



#74
 N-ethyl acetate
 Concen: 0.220 ug/l
 RT: 10.01 min Scan# 2774
 Delta R.T. 0.01 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion: 43 Resp: 2148
 Ion Ratio Lower Upper
 43 100
 70 22.1 35.2 52.8#
 55 19.1 20.5 30.7#
 61 0.0 20.0 30.0#





#76

1,2,3-Trichloropropane

Concen: 20.242 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.19 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampleId :





GW-5

Tgt Ion: 75 Resp: 157753

Ion Ratio Lower Upper

75 100

77 1.3 20.1 60.2#



#81

trans-1,4-Dichloro-2-butene

Concen: 68.161 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.21 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19




Tgt Ion: 75 Resp: 157753


Ion Ratio Lower Upper

75 100

53 0.0 0.0 0.0#

89 0.0 0.0 0.0









#82
 4-Chlorotoluene
 Concen: 0.207 ug/l
 RT: 10.76 min Scan# 3008
 Delta R.T. 0.01 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5


Tgt Ion: 91 Resp: 3864
 Ion Ratio Lower Upper
 91 100
 126 26.5 15.0 45.1



#87
 1,3-Dichlorobenzene
 Concen: 0.281 ug/l
 RT: 11.40 min Scan# 3208
 Delta R.T. 0.01 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion: 146 Resp: 3107
 Ion Ratio Lower Upper
 146 100
 111 39.1 19.9 59.6
 148 59.6 31.9 95.5









#88
 1,4-Dichlorobenzene
 Concen: 0.278 ug/l
 RT: 11.40 min Scan# 3208
 Delta R.T. -0.08 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5


Tgt Ion:146 Resp: 3107
 Ion Ratio Lower Upper
 146 100
 111 39.1 19.2 57.6
 148 59.6 31.9 95.9



#89
 n-Butylbenzene
 Concen: 0.266 ug/l
 RT: 11.88 min Scan# 3356
 Delta R.T. 0.01 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion: 91 Resp: 4918
 Ion Ratio Lower Upper
 91 100
 92 44.0 26.3 78.9
 134 16.4 12.9 38.7









#91
 1,2-Dichlorobenzene
 Concen: 0.261 ug/l
 RT: 11.87 min Scan# 3352
 Delta R.T. 0.01 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19




Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5


Tgt Ion:146 Resp: 2766
 Ion Ratio Lower Upper
 146 100
 111 34.9 20.4 61.2
 148 55.0 32.3 96.8



#93
 1,2,4-Trichlorobenzene
 Concen: 1.377 ug/l
 RT: 13.50 min Scan# 3861
 Delta R.T. 0.02 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Tgt Ion:180 Resp: 2026
 Ion Ratio Lower Upper
 180 100
 182 85.4 47.4 142.2
 145 20.0 14.1 42.3





#94

Hexachlorobutadiene

Concen: 0.357 ug/l

RT: 13.68 min Scan# 3915

Delta R.T. 0.00 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19

Instrument :

MSVOA_U

ClientSampled :

GW-5


Tgt Ion: 225 Resp: 1478

Ion Ratio Lower Upper

225 100

223 62.3 31.8 95.3




227 73.0 32.6 97.8



Abundance Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 2.136 ug/l

RT: 13.75 min Scan# 3938

Delta R.T. 0.03 min

Lab File: VU034028.D

Acq: 28 Aug 2019 17:19


Tgt Ion: 128 Resp: 7751

Ion Ratio Lower Upper

128 100

127 12.1 10.2 15.4



129 7.1 8.4 12.6#




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 0.596 ug/l
 RT: 13.98 min Scan# 4010
 Delta R.T. 0.01 min
 Lab File: VU034028.D
 Acq: 28 Aug 2019 17:19

Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5

Tgt Ion	Ratio	Lower	Upper
180	100		
182	85.4	48.2	144.6
145	21.1	15.5	46.5

