

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U082919\
 Data File : VU034028.D
 Acq On : 28 Aug 2019 17:19
 Operator : JC/SP
 Sample : K4592-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GW-5

Quant Time: Aug 29 04:49:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	477327	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	756278	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	708578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	345777	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	291447	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
35) Dibromofluoromethane	4.86	113	243792	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
50) Toluene-d8	7.55	98	940063	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
62) 4-Bromofluorobenzene	10.29	95	329094	43.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.14%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.82	59	6084	4.314	ug/l #	85
16) Acetone	2.31	43	10946	3.507	ug/l	93
17) Carbon Disulfide	2.47	76	11785	0.779	ug/l	96
95) Naphthalene	13.75	128	7751	2.136	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082919\
Data File : VU034028.D
Acq On : 28 Aug 2019 17:19
Operator : JC/SP
Sample : K4592-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GW-5

Quant Time: Aug 29 04:49:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

