

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082919\
 Data File : VU034032.D
 Acq On : 28 Aug 2019 18:51
 Operator : JC/SP
 Sample : PB122589TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB122589TB

Manual Integrations
 APPROVED

MMDadoda
 8/29/2019 8:47:36 AM

Quant Time: Aug 29 04:59:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	478812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	761070	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	724775	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	347095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	292379	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
35) Dibromofluoromethane	4.86	113	242635	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
50) Toluene-d8	7.55	98	961308	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
62) 4-Bromofluorobenzene	10.29	95	336785	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
Target Compounds						
20) Methylene Chloride	2.69	84	10495	1.771	ug/l	99
43) Isopropyl Acetate	5.53	43	100783	8.218	ug/l	99
95) Naphthalene	13.75	128	13273m	2.381	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082919\
Data File : VU034032.D
Acq On : 28 Aug 2019 18:51
Operator : JC/SP
Sample : PB122589TB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB122589TB

Manual Integrations
APPROVED

MMDadoda
8/29/2019 8:47:36 AM

Quant Time: Aug 29 04:59:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

