

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082919\
 Data File : VU034037.D
 Acq On : 28 Aug 2019 20:48
 Operator : JC/SP
 Sample : K4525-26
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T9-12-13-A1-(1.9)

Quant Time: Aug 29 05:17:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	472976	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	768005	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	739608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	364612	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	296639	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
35) Dibromofluoromethane	4.86	113	248199	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
50) Toluene-d8	7.55	98	969407	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	10.29	95	342962	44.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.40%	
Target Compounds						
20) Methylene Chloride	2.69	84	14862	2.539	ug/l	96
43) Isopropyl Acetate	5.53	43	96643	7.809	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082919\
Data File : VU034037.D
Acq On : 28 Aug 2019 20:48
Operator : JC/SP
Sample : K4525-26
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T9-12-13-A1-(1.9)

Quant Time: Aug 29 05:17:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

