

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082919\
 Data File : VU034045.D
 Acq On : 28 Aug 2019 23:54
 Operator : JC/SP
 Sample : K4525-35
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T9-12-13-B1-(1.9)

Quant Time: Aug 29 05:30:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	466482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	754880	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	738521	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	364662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	295488	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
35) Dibromofluoromethane	4.86	113	247135	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	7.55	98	976410	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	10.29	95	343966	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
16) Acetone	2.31	43	25282	8.290	ug/l	95
20) Methylene Chloride	2.69	84	42865	7.424	ug/l	98
43) Isopropyl Acetate	5.53	43	102085	8.392	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082919\
Data File : VU034045.D
Acq On : 28 Aug 2019 23:54
Operator : JC/SP
Sample : K4525-35
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T9-12-13-B1-(1.9)

Quant Time: Aug 29 05:30:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.96 min Scan# 1204

Delta R.T. 0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54


Instrument :
MSVOA_U
ClientSampleId :
T9-12-13-B1-(1.9)

Tot Ion:168 Resp: 466482




Ion Ratio Lower Upper

168 100

99 54.0 42.3 63.5



Abundance Ion 167.90 (167.60 to 168.60): VU034018.D (-1185) (-)



#16

Acetone

Concen: 8.290 ug/l

RT: 2.31 min Scan# 380

Delta R.T. 0.00 min

Lab File: VU034045.D


Acq: 28 Aug 2019 23:54

Tgt Ion: 43 Resp: 25282



Ion Ratio Lower Upper


43 100

58 36.4 26.8 40.2



Abundance Ion 43.00 (42.70 to 43.70): VU034018.D (-358) (-)





#20

Methylene Chloride

Concen: 7.424 ug/l

RT: 2.69 min Scan# 497

Delta R.T. 0.01 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-B1-(1.9)

Tot Ion: 84 Resp: 42865


Ion Ratio Lower Upper

84 100

49 126.3 100.2 150.4

51 40.0 30.8 46.2

86 67.7 52.5 78.7






Abundance Ion 83.90 (83.60 to 84.60): VU0

Ion 48.90 (48.60 to 49.60): VU0

Ion 50.90 (50.60 to 51.60): VU0

Ion 85.90 (85.60 to 86.60): VU0



#33

1,2-Dichloroethane-d4

Concen: 49.436 ug/l

RT: 5.29 min Scan# 1306

Delta R.T. 0.00 min

Lab File: VU034045.D


Acq: 28 Aug 2019 23:54

Tgt Ion: 65 Resp: 295488

Ion Ratio Lower Upper



65 100

67 55.2 0.0 110.6



Abundance Ion 64.90 (64.60 to 65.60): VU0

Ion 66.90 (66.60 to 67.60): VU0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.87 min Scan# 1486

Delta R.T. 0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-B1-(1.9)


Tot Ion:114 Resp: 754880

Ion Ratio Lower Upper

114 100

63 21.4 0.0 42.0




88 15.8 0.0 32.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VU0

Ion 87.90 (87.60 to 88.60): VU0



#35

Dibromofluoromethane

Concen: 47.201 ug/l

RT: 4.86 min Scan# 1173

Delta R.T. 0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54


Tgt Ion:113 Resp: 247135

Ion Ratio Lower Upper

113 100

111 101.6 81.7 122.5



192 21.8 17.1 25.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 8.392 ug/l

RT: 5.53 min Scan# 1381

Delta R.T. 0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-B1-(1.9)


Tot Ion: 43 Resp: 102085

Ion Ratio Lower Upper

43 100

61 20.8 17.2 25.8

87 12.6 10.8 16.2




Abundance Ion 43.00 (42.70 to 43.70): VU034018.D (-1365) (-)

Ion 61.00 (60.70 to 61.70): VU034018.D (-1365) (-)


Ion 87.00 (86.70 to 87.70): VU034018.D (-1365) (-)

5.53

Time-->



Time-->



#50

Toluene-d8

Concen: 49.805 ug/l

RT: 7.55 min Scan# 2008

Delta R.T. -0.00 min

Lab File: VU034045.D


Acq: 28 Aug 2019 23:54

Tgt Ion: 98 Resp: 976410

Ion Ratio Lower Upper

98 100

100 64.0 50.3 75.5




Abundance Ion 98.00 (97.70 to 98.70): VU034018.D (-1991) (-)


Ion 100.00 (99.70 to 100.70): VU034018.D (-1991) (-)

7.55

Time-->



Time-->



#62

4-Bromofluorobenzene

Concen: 45.104 ug/l

RT: 10.29 min Scan# 2861

Delta R.T. -0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-B1-(1.9)


Tot Ion: 95 Resp: 343966

Ion Ratio Lower Upper

95 100

174 83.2 0.0 168.8

176 79.2 0.0 161.8






Abundance Ion 94.90 (94.60 to 95.60): VU034018.D (-2848) (-)

Ion 173.80 (173.50 to 174.50): VU034018.D (-2848) (-)

Ion 175.80 (175.50 to 176.50): VU034018.D (-2848) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.07 min Scan# 2482

Delta R.T. -0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54


Tgt Ion: 117 Resp: 738521

Ion Ratio Lower Upper

117 100

82 57.0 45.1 67.7

119 32.0 25.4 38.0






Abundance Ion 116.90 (116.60 to 117.60): VU034018.D (-2468) (-)

Ion 82.00 (81.70 to 82.70): VU034018.D (-2468) (-)

Ion 118.90 (118.60 to 119.60): VU034018.D (-2468) (-)

Time-->





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 11.47 min Scan# 3227

Delta R.T. -0.00 min

Lab File: VU034045.D

Acq: 28 Aug 2019 23:54

Instrument :

MSVOA_U

ClientSampleId :

T9-12-13-B1-(1.9)


Tot Ion:152 Resp: 364662

Ion Ratio Lower Upper

152 100

115 57.1 41.1 123.3

150 155.8 0.0 348.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

