

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082922\
 Data File : VU050475.D
 Acq On : 29 Aug 2022 11:21
 Operator : SY/MD
 Sample : VU0829MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK170

Quant Time: Aug 30 02:12:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM082422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 30 02:11:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	242658	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	237531	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	80155	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	123957	43.202	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.400%	
7) Chloroethane-d5	1.900	69	102787	48.414	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.820%	
11) 1,1-Dichloroethene-d2	2.562	63	147634	36.409	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	72.820%	
21) 2-Butanone-d5	4.626	46	220755	117.545	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	117.540%	
24) Chloroform-d	5.060	84	218754	47.494	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.980%	
26) 1,2-Dichloroethane-d4	5.700	65	140497	49.081	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.160%	
32) Benzene-d6	5.726	84	424982	48.422	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.840%	
36) 1,2-Dichloropropane-d6	6.690	67	146955	50.543	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.080%	
41) Toluene-d8	7.899	98	334635	43.574	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.140%	
43) trans-1,3-Dichloroprop...	8.179	79	53342	43.981	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.960%	
47) 2-Hexanone-d5	8.636	63	105275	97.567	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	97.570%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	223927	49.933	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.860%	
66) 1,2-Dichlorobenzene-d4	12.195	152	102883	53.732	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	107.460%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082922\
Data File : VU050475.D
Acq On : 29 Aug 2022 11:21
Operator : SY/MD
Sample : VU0829MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK170

Quant Time: Aug 30 02:12:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM082422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 30 02:11:49 2022
Response via : Initial Calibration

