

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082923\
 Data File : VU054979.D
 Acq On : 30 Aug 2023 01:14
 Operator : MD/SY
 Sample : 04209-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH486

Quant Time: Aug 30 01:43:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 30 01:18:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	159874	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	156102	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	77278	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	24380	4.832	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.600%	
7) Chloroethane-d5	1.907	69	31282	4.799	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.000%	
11) 1,1-Dichloroethene-d2	2.563	65	14267	5.003	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	100.000%	
20) 2-Butanone-d5	4.628	46	163588	60.060	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.120%	
24) Chloroform-d	5.062	84	92316	4.932	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.600%	
26) 1,2-Dichloroethane-d4	5.702	65	52075	5.228	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.600%	
32) Benzene-d6	5.727	84	166193	4.736	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.800%	
36) 1,2-Dichloropropane-d6	6.689	67	60345	4.890	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.800%	
41) Toluene-d8	7.898	98	145772	4.338	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.800%	
43) trans-1,3-Dichloroprop...	8.181	79	16507	4.064	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.200%	
46) 2-Hexanone-d5	8.634	63	101092	52.442	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	104.880%	
56) 1,1,2,2-Tetrachloroeth...	10.756	84	56589	5.347	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.000%	
66) 1,2-Dichlorobenzene-d4	12.193	152	61738	5.004	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.000%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.518	50	5595	0.541	ug/L	90
22) cis-1,2-Dichloroethene	4.666	96	9283	0.987	ug/L	91
31) Carbon tetrachloride	5.518	117	4067	0.308	ug/L	98
34) Trichloroethene	6.544	95	18385	1.801	ug/L	94
47) Tetrachloroethene	8.553	164	8047	1.097	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082923\
Data File : VU054979.D
Acq On : 30 Aug 2023 01:14
Operator : MD/SY
Sample : 04209-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH486

Quant Time: Aug 30 01:43:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 30 01:18:16 2023
Response via : Initial Calibration









#3
Chloromethane
Concen: 0.541 ug/L
RT: 1.518 min Scan# 71
Delta R.T. 0.000 min
Lab File: VU054979.D
Acq: 30 Aug 2023 01:14




Instrument :
MSVOA_U
ClientSampleId :
BH486


Tgt Ion: 50 Resp: 5595
Ion Ratio Lower Upper
50 100
52 27.8 23.4 43.6



#22
cis-1,2-Dichloroethene
Concen: 0.987 ug/L
RT: 4.666 min Scan# 1050
Delta R.T. 0.003 min
Lab File: VU054979.D
Acq: 30 Aug 2023 01:14

Tgt Ion: 96 Resp: 9283
Ion Ratio Lower Upper
96 100
61 119.3 88.4 164.2
98 51.9 44.0 81.6









#31
Carbon tetrachloride
Concen: 0.308 ug/L
RT: 5.518 min Scan# 11
Delta R.T. -0.003 min
Lab File: VU054979.D
Acq: 30 Aug 2023 01:14




Instrument :
MSVOA_U
ClientSampleId :
BH486


Tgt Ion:117 Resp: 4067
Ion Ratio Lower Upper
117 100
119 93.8 76.6 115.0



#34
Trichloroethene
Concen: 1.801 ug/L
RT: 6.544 min Scan# 1634
Delta R.T. 0.003 min
Lab File: VU054979.D
Acq: 30 Aug 2023 01:14


Tgt Ion: 95 Resp: 18385
Ion Ratio Lower Upper
95 100
97 66.8 44.1 81.9
132 93.6 63.6 118.2
130 99.4 63.6 118.0





#47
 Tetrachloroethene
 Concen: 1.097 ug/L
 RT: 8.553 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VU054979.D
 Acq: 30 Aug 2023 01:14

Instrument :
 MSVOA_U
 ClientSampleId :
 BH486



Tgt Ion:	164	Resp:	8047
Ion	Ratio	Lower	Upper
164	100		
129	103.4	72.8	135.2
131	94.6	70.7	131.3
166	128.9	95.9	178.1

