

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083018\
 Data File : VU026471.D
 Acq On : 31 Aug 2018 01:05
 Operator : MD/SY
 Sample : J4694-10 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEE28

Quant Time: Aug 31 06:41:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 31 05:38:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	151110	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	141225	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	61745	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	40308	43.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.10%
7) Chloroethane-d5	1.68	69	37033	43.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.60%
11) 1,1-Dichloroethene-d2	2.27	63	58921	27.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.90%#
21) 2-Butanone-d5	4.18	46	66221	91.13	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.13%
24) Chloroform-d	4.65	84	75005	42.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.84%
26) 1,2-Dichloroethane-d4	5.31	65	52053	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.44%
32) Benzene-d6	5.34	84	150869	46.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.70%
36) 1,2-Dichloropropane-d6	6.33	67	52334	46.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.60%
41) Toluene-d8	7.57	98	137101	43.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.26%
43) trans-1,3-Dichloropropene-	7.85	79	22016	43.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.24%
47) 2-Hexanone-d5	8.31	63	47740	93.09	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.09%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	72914	44.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	54050	48.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.50%


Target Compounds					Ovalue
46) Tetrachloroethene	8.23	164	2872	3.46 ug/L	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083018\
Data File : VU026471.D
Acq On : 31 Aug 2018 01:05
Operator : MD/SY
Sample : J4694-10 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEE28

Quant Time: Aug 31 06:41:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 31 05:38:13 2018
Response via : Initial Calibration





#46
Tetrachloroethene
Concen: 3.46 ug/L
RT: 8.23 min Scan# 2376
Delta R.T. 0.00 min
Lab File: VU026471.D
Acq: 31 Aug 2018 01:05

Instrument :
MSVOA_U
ClientSampleId :
BEE28

Tot Ion:164 Resp: 2872
Ion Ratio Lower Upper
164 100
129 85.0 67.3 125.1
131 80.8 65.3 121.3
166 111.0 90.9 168.7

