

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083019\
 Data File : VU034062.D
 Acq On : 29 Aug 2019 15:17
 Operator : JC/SP
 Sample : PB122636TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB122636TB

Quant Time: Aug 29 16:06:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	468072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	735916	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	733120	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	355828	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	294653	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
35) Dibromofluoromethane	4.86	113	241445	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
50) Toluene-d8	7.55	98	971269	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	10.29	95	339013	45.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.20%	
Target Compounds						
43) Isopropyl Acetate	5.53	43	95248	8.032	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083019\
Data File : VU034062.D
Acq On : 29 Aug 2019 15:17
Operator : JC/SP
Sample : PB122636TB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB122636TB

Quant Time: Aug 29 16:06:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

