

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083019\
 Data File : VU034063.D
 Acq On : 29 Aug 2019 15:40
 Operator : JC/SP
 Sample : K4596-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB-3-082819

Quant Time: Aug 30 04:56:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	455470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	723298	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	696228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	345013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	292057	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	4.86	113	238041	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	7.55	98	912710	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	10.29	95	322163	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
Target Compounds						
16) Acetone	2.31	43	26901	9.034	ug/l	99
20) Methylene Chloride	2.69	84	19628	3.482	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083019\
Data File : VU034063.D
Acq On : 29 Aug 2019 15:40
Operator : JC/SP
Sample : K4596-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB-3-082819

Quant Time: Aug 30 04:56:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

