

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083019\
 Data File : VU034065.D
 Acq On : 29 Aug 2019 16:27
 Operator : JC/SP
 Sample : K4596-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB-5-082819

Quant Time: Aug 30 05:01:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	458835	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	740262	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	703837	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	351136	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	291767	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
35) Dibromofluoromethane	4.86	113	241309	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
50) Toluene-d8	7.55	98	931731	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	10.29	95	323593	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	
Target Compounds						
16) Acetone	2.31	43	31063	10.355	ug/l	99
20) Methylene Chloride	2.69	84	19627	3.456	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083019\
Data File : VU034065.D
Acq On : 29 Aug 2019 16:27
Operator : JC/SP
Sample : K4596-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB-5-082819

Quant Time: Aug 30 05:01:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

