

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083019\
 Data File : VU034067.D
 Acq On : 29 Aug 2019 17:13
 Operator : JC/SP
 Sample : K4595-05ME
 Misc : 6.05g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S503-K2-9.5-10.0-082819ME

Quant Time: Aug 30 07:29:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.95	168	477493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	757440	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	729094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.47	152	373863	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.28	65	294337	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
35) Dibromofluoromethane	4.86	113	238918	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
50) Toluene-d8	7.55	98	940198	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
62) 4-Bromofluorobenzene	10.30	95	356845	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.20	96	15747	2.704	ug/l	86
44) Trichloroethene	6.16	130	30908	5.430	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083019\
Data File : VU034067.D
Acq On : 29 Aug 2019 17:13
Operator : JC/SP
Sample : K4595-05ME
Misc : 6.05g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S503-K2-9.5-10.0-082819ME

Quant Time: Aug 30 07:29:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

