

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU083019\
 Data File : VU034081.D
 Acq On : 29 Aug 2019 22:38
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 30 05:43:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.503	0.514	-2.2	90	0.00
3 P	Chloromethane	0.749	0.597	20.3#	81	0.00
4 C	Vinyl Chloride	0.669	0.640	4.3#	88	0.00
5 T	Bromomethane	0.351	0.225	35.9#	62	0.01
6 T	Chloroethane	0.409	0.408	0.2	95	0.00
7 T	Trichlorofluoromethane	0.775	0.774	0.1	92	0.00
8 T	Diethyl Ether	0.332	0.335	-0.9	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.478	2.6	91	0.00
10 T	Methyl Iodide	0.697	0.249	64.3#	32#	0.00
11 T	Tert butyl alcohol	0.148	0.131	11.5	83	-0.01
12 CM	1,1-Dichloroethene	0.501	0.500	0.2#	93	0.00
13 T	Acrolein	0.062	0.050	19.4	78	0.00
14 T	Allyl chloride	0.941	0.822	12.6	81	0.00
15 T	Acrylonitrile	0.339	0.349	-2.9	93	0.00
16 T	Acetone	0.327	0.285	12.8	82	0.00
17 T	Carbon Disulfide	1.584	1.481	6.5	89	0.00
18 T	Methyl Acetate	0.845	0.816	3.4	96	0.00
19 T	Methyl tert-butyl Ether	1.640	1.723	-5.1	94	0.00
20 T	Methylene Chloride	0.619	0.609	1.6	95	0.00
21 T	trans-1,2-Dichloroethene	0.543	0.542	0.2	92	0.00
22 T	Diisopropyl ether	1.725	1.810	-4.9	95	0.00
23 T	Vinyl Acetate	1.481	1.564	-5.6	93	0.00
24 P	1,1-Dichloroethane	1.018	1.035	-1.7	94	0.00
25 T	2-Butanone	0.455	0.456	-0.2	88	0.00
26 T	2,2-Dichloropropane	0.828	0.735	11.2	82	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.613	-0.5	93	0.00
28 T	Bromochloromethane	0.502	0.513	-2.2	95	0.00
29 T	Tetrahydrofuran	0.280	0.296	-5.7	93	0.00
30 C	Chloroform	1.016	0.995	2.1#	94	0.00
31 T	Cyclohexane	0.957	0.910	4.9	90	0.00
32 T	1,1,1-Trichloroethane	0.832	0.834	-0.2	93	0.00
33 S	1,2-Dichloroethane-d4	0.641	0.647	-0.9	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.347	0.341	1.7	96	0.00
36 T	1,1-Dichloropropene	0.473	0.470	0.6	94	0.00
37 T	Ethyl Acetate	0.530	0.525	0.9	90	0.00
38 T	Carbon Tetrachloride	0.457	0.447	2.2	92	0.00
39 T	Methylcyclohexane	0.557	0.542	2.7	89	0.00
40 TM	Benzene	1.431	1.448	-1.2	95	0.00
41 T	Methacrylonitrile	0.272	0.286	-5.1	93	0.00
42 TM	1,2-Dichloroethane	0.494	0.491	0.6	93	0.00
43 T	Isopropyl Acetate	0.806	0.811	-0.6	92	0.00
44 TM	Trichloroethene	0.376	0.390	-3.7	99	0.00
45 C	1,2-Dichloropropane	0.394	0.400	-1.5#	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU083019\
 Data File : VU034081.D
 Acq On : 29 Aug 2019 22:38
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 30 05:43:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 29 04:11:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.254	0.254	0.0	94	0.00
47 T	Bromodichloromethane	0.496	0.484	2.4	92	0.00
48 T	Methyl methacrylate	0.389	0.396	-1.8	90	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	81	0.00
50 S	Toluene-d8	1.299	1.342	-3.3	96	0.00
51 T	4-Methyl-2-Pentanone	0.556	0.551	0.9	91	0.00
52 CM	Toluene	0.881	0.917	-4.1#	96	0.00
53 T	t-1,3-Dichloropropene	0.551	0.527	4.4	88	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.592	1.7	90	0.00
55 T	1,1,2-Trichloroethane	0.373	0.367	1.6	93	0.00
56 T	Ethyl methacrylate	0.557	0.557	0.0	90	0.00
57 T	1,3-Dichloropropane	0.633	0.633	0.0	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.206	0.200	2.9	78	0.00
59 T	2-Hexanone	0.420	0.419	0.2	89	0.00
60 T	Dibromochloromethane	0.400	0.388	3.0	91	0.00
61 T	1,2-Dibromoethane	0.393	0.382	2.8	92	0.00
62 S	4-Bromofluorobenzene	0.505	0.503	0.4	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.336	0.350	-4.2	99	0.00
65 PM	Chlorobenzene	1.029	1.021	0.8	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.352	5.6	91	0.00
67 C	Ethyl Benzene	1.744	1.782	-2.2#	94	0.00
68 T	m/p-Xylenes	0.663	0.682	-2.9	94	0.00
69 T	o-Xylene	0.636	0.652	-2.5	93	0.00
70 T	Styrene	1.091	1.153	-5.7	94	0.00
71 P	Bromoform	0.324	0.297	8.3	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	3.358	3.254	3.1	93	0.00
74 T	N-amyl acetate	1.413	1.395	1.3	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.285	1.174	8.6	93	0.00
76 T	1,2,3-Trichloropropane	1.127	1.001	11.2	86	0.00
77 T	Bromobenzene	0.903	0.853	5.5	94	0.00
78 T	n-propylbenzene	3.911	3.819	2.4	93	0.00
79 T	2-Chlorotoluene	2.346	2.274	3.1	94	0.00
80 T	1,3,5-Trimethylbenzene	2.847	2.810	1.3	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.335	0.238	29.0#	71	0.00
82 T	4-Chlorotoluene	2.698	2.634	2.4	95	0.00
83 T	tert-Butylbenzene	2.802	2.633	6.0	90	0.00
84 T	1,2,4-Trimethylbenzene	2.827	2.844	-0.6	93	0.00
85 T	sec-Butylbenzene	3.359	3.223	4.0	91	0.00
86 T	p-Isopropyltoluene	3.038	2.979	1.9	92	0.00
87 T	1,3-Dichlorobenzene	1.599	1.548	3.2	96	0.00
88 T	1,4-Dichlorobenzene	1.618	1.540	4.8	96	0.00
89 T	n-Butylbenzene	2.678	2.663	0.6	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU083019\
Data File : VU034081.D
Acq On : 29 Aug 2019 22:38
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Aug 30 05:43:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 29 04:11:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.525	0.403	23.2#	76	0.00
91 T	1,2-Dichlorobenzene	1.532	1.505	1.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.256	0.232	9.4	87	0.00
93 T	1,2,4-Trichlorobenzene	0.937	1.027	-9.6	94	0.00
94 T	Hexachlorobutadiene	0.599	0.477	20.4#	78	0.00
95 T	Naphthalene	2.584	2.988	-15.6	97	0.00
96 T	1,2,3-Trichlorobenzene	0.981	1.015	-3.5	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6