


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU083019\
Data File : VU034051.D
Acq On : 29 Aug 2019 10:49
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U082919W.M
Title : SW846 8260
Last Update : Thu Aug 29 04:11:12 2019



AutoFind: Scans 2861, 2862, 2863; Background Corrected with Scan 2848

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	18602	PASS
75	95	30	60	48.3	50626	PASS
95	95	100	100	100.0	104781	PASS
96	95	5	9	6.9	7224	PASS
173	174	0.00	2	0.4	315	PASS
174	95	50	100	85.6	89658	PASS
175	174	5	9	7.3	6510	PASS
176	174	95	101	97.6	87504	PASS
177	176	5	9	6.4	5633	PASS