

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083023\
 Data File : VU054994.D
 Acq On : 30 Aug 2023 13:12
 Operator : MD/SY
 Sample : 04209-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH480

Quant Time: Aug 31 02:35:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Aug 31 02:34:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	163487	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	157411	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	79930	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	19639	3.806	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.200%
7) Chloroethane-d5	1.907	69	28098	4.215	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.563	65	12802	4.390	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	4.628	46	172821	62.048	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.100%
24) Chloroform-d	5.062	84	92385	4.827	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.600%
26) 1,2-Dichloroethane-d4	5.702	65	54590	5.360	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%
32) Benzene-d6	5.727	84	161973	4.578	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.689	67	61058	4.906	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.898	98	145285	4.287	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	8.181	79	18327	4.475	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.634	63	116407	59.885	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.760%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	56535	5.298	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.000%
66) 1,2-Dichlorobenzene-d4	12.193	152	63823	5.001	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds						
					Qvalue	
3) Chloromethane	1.515	50	2277	0.215	ug/L	# 59
22) cis-1,2-Dichloroethene	4.666	96	16238	1.688	ug/L	98
34) Trichloroethene	6.541	95	32144	3.123	ug/L	96
47) Tetrachloroethene	8.550	164	9257	1.252	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083023\
Data File : VU054994.D
Acq On : 30 Aug 2023 13:12
Operator : MD/SY
Sample : 04209-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH480

Quant Time: Aug 31 02:35:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Aug 31 02:34:15 2023
Response via : Initial Calibration

