

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083023\
 Data File : VU055009.D
 Acq On : 30 Aug 2023 20:35
 Operator : MD/SY
 Sample : VIBLK126
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK126

Quant Time: Aug 31 02:40:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Aug 31 02:34:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	185917	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	187665	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	96028	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0d	0.000	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.000%#
7) Chloroethane-d5	1.907	69	869	0.115	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	2.200%#
11) 1,1-Dichloroethene-d2	0.000	65	0d	0.000	ug/L	
Spiked Amount	5.000	Range	60 - 125	Recovery	=	0.000%#
20) 2-Butanone-d5	4.628	46	190964	60.290	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.580%
24) Chloroform-d	5.062	84	13525	0.621	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	12.400%#
26) 1,2-Dichloroethane-d4	5.708	65	10685	0.922	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	18.400%#
32) Benzene-d6	5.724	84	4391	0.104	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	2.000%#
36) 1,2-Dichloropropane-d6	6.685	67	7393	0.498	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	10.000%#
41) Toluene-d8	7.901	98	2770	0.069	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	1.400%#
43) trans-1,3-Dichloroprop...	8.181	79	2647	0.542	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	10.800%#
46) 2-Hexanone-d5	8.634	63	113179	48.838	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.680%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	46373	3.645	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.800%
66) 1,2-Dichlorobenzene-d4	12.193	152	19749	1.288	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	25.800%#
Target Compounds						
					Qvalue	
16) Methylene chloride	3.039	84	42680	3.309	ug/L	98
22) cis-1,2-Dichloroethene	4.669	96	5043	0.461	ug/L	86
34) Trichloroethene	6.541	95	16444	1.340	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083023\
Data File : VU055009.D
Acq On : 30 Aug 2023 20:35
Operator : MD/SY
Sample : VIBLK126
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK126

Quant Time: Aug 31 02:40:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Aug 31 02:34:15 2023
Response via : Initial Calibration

