

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083024\
 Data File : VU060573.D
 Acq On : 30 Aug 2024 15:20
 Operator : MD/SY
 Sample : P3768-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OW-03B-51.5-082724-VC-FD

Quant Time: Aug 31 01:02:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	5396	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	8106	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2338	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	3616	0.412	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	82.000%
4) 1,2-Dichloroethane-d4	5.685	65	2155	0.463	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
7) 1,2-Dichloropropane-d6	6.687	67	2442	0.485	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	96.000%
8) Toluene-d8	7.888	98	5612	0.539	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	108.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1973	0.447	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	90.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083024\
Data File : VU060573.D
Acq On : 30 Aug 2024 15:20
Operator : MD/SY
Sample : P3768-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OW-03B-51.5-082724-VC-FD

Quant Time: Aug 31 01:02:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration

