

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083118\
 Data File : VU026503.D
 Acq On : 31 Aug 2018 16:35
 Operator : MD/SY
 Sample : J4691-18
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEE15

Quant Time: Sep 01 04:42:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 01 01:02:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	148429	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	139779	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	62828	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	40831	44.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.80%
7) Chloroethane-d5	1.68	69	38695	46.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.18%
11) 1,1-Dichloroethene-d2	2.27	63	58873	28.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.86%#
21) 2-Butanone-d5	4.18	46	73616	103.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.14%
24) Chloroform-d	4.65	84	82640	47.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.16%
26) 1,2-Dichloroethane-d4	5.31	65	58112	50.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.50%
32) Benzene-d6	5.34	84	164174	50.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.92%
36) 1,2-Dichloropropane-d6	6.33	67	57046	51.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.08%
41) Toluene-d8	7.57	98	150447	48.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.76%
43) trans-1,3-Dichloropropene-	7.85	79	22664	44.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.68%
47) 2-Hexanone-d5	8.31	63	55093	108.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.54%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	79545	48.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.56%
64) 1,2-Dichlorobenzene-d4	11.86	152	60452	53.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.06%

Target Compounds

						Ovalue
3) Chloromethane	1.33	50	16440	11.50	ug/L	98
13) Acetone	2.32	43	26697	43.20	ug/L	98
14) Carbon disulfide	2.48	76	3836	1.28	ug/L	98
20) cis-1,2-Dichloroethene	4.23	96	10376	9.67	ug/L	98
22) 2-Butanone	4.27	43	10946	13.18	ug/L	100
33) Benzene	5.39	78	15674	3.91	ug/L	100
34) Trichloroethene	6.19	95	7907	7.87	ug/L	95
46) Tetrachloroethene	8.23	164	227783	277.18	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083118\
Data File : VU026503.D
Acq On : 31 Aug 2018 16:35
Operator : MD/SY
Sample : J4691-18
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEE15

Quant Time: Sep 01 04:42:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 01 01:02:22 2018
Response via : Initial Calibration

