

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083118\
 Data File : VU026507.D
 Acq On : 31 Aug 2018 18:08
 Operator : MD/SY
 Sample : J4691-10DL 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEE07DL

Quant Time: Sep 04 09:58:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 01 01:02:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	145980	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	137019	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	58498	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	37463	41.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.84%
7) Chloroethane-d5	1.68	69	36780	45.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.06%
11) 1,1-Dichloroethene-d2	2.27	63	61738	30.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.62%
21) 2-Butanone-d5	4.18	46	71497	101.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.85%
24) Chloroform-d	4.65	84	71053	41.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.18%
26) 1,2-Dichloroethane-d4	5.31	65	55521	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.64%
32) Benzene-d6	5.35	84	157957	50.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.04%
36) 1,2-Dichloropropane-d6	6.33	67	55532	51.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.38%
41) Toluene-d8	7.57	98	141312	46.35	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.70%
43) trans-1,3-Dichloropropene-	7.85	79	23990	48.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.84%
47) 2-Hexanone-d5	8.31	63	52443	105.40	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.40%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	79534	49.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	58662	55.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.54%

Target Compounds						Ovalue
34) Trichloroethene	6.19	95	3620	3.67	ug/L	95
46) Tetrachloroethene	8.23	164	21827	27.10	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083118\
Data File : VU026507.D
Acq On : 31 Aug 2018 18:08
Operator : MD/SY
Sample : J4691-10DL 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEE07DL

Quant Time: Sep 04 09:58:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 01 01:02:22 2018
Response via : Initial Calibration

