

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU083118\
 Data File : VU026488.D
 Acq On : 31 Aug 2018 10:31
 Operator : MD/SY
 Sample : VU0831MBL01
 Misc : 5.00uL/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK57

Quant Time: Sep 01 01:11:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 01 01:02:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	155536	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	145974	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	61804	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	43807	45.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.92%
7) Chloroethane-d5	1.68	69	40426	46.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.90%
11) 1,1-Dichloroethene-d2	2.27	63	71253	32.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.66%
21) 2-Butanone-d5	4.18	46	74153	99.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.14%
24) Chloroform-d	4.65	84	88453	48.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	5.31	65	59459	49.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.14%
32) Benzene-d6	5.34	84	174126	51.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.50%
36) 1,2-Dichloropropane-d6	6.33	67	59721	51.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.34%
41) Toluene-d8	7.57	98	160674	49.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.94%
43) trans-1,3-Dichloropropene-	7.85	79	26222	49.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.36%
47) 2-Hexanone-d5	8.31	63	52836	99.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.67%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	82254	48.30	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	62647	55.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU083118\
Data File : VU026488.D
Acq On : 31 Aug 2018 10:31
Operator : MD/SY
Sample : VU0831MBL01
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK57

Quant Time: Sep 01 01:11:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 01 01:02:22 2018
Response via : Initial Calibration

