


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083121\
 Data File : VU044533.D
 Acq On : 31 Aug 2021 16:11
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U083121DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Wed Sep 01 01:29:46 2021



AutoFind: Averaged scan 2879 to 2916; Bkg corrected with scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	2684	PASS
75	95	30	60	53.2	7397	PASS
95	95	100	100	100.0	13901	PASS
96	95	5	9	6.8	949	PASS
173	174	0.00	2	1.2	143	PASS
174	95	50	100	87.0	12090	PASS
175	174	5	9	7.4	890	PASS
176	174	95	101	95.5	11541	PASS
177	176	5	9	6.3	726	PASS