

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083123\
 Data File : VU055014.D
 Acq On : 31 Aug 2023 12:18
 Operator : MD/SY
 Sample : 04230-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH497

Quant Time: Sep 01 04:59:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 01 04:57:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	182366	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	169837	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	83398	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	18097	3.144	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.800%
7) Chloroethane-d5	1.907	69	26284	3.535	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.600%
11) 1,1-Dichloroethene-d2	2.560	65	11028	3.390	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.800%
20) 2-Butanone-d5	4.628	46	144602	46.542	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.080%
24) Chloroform-d	5.058	84	87236	4.086	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
26) 1,2-Dichloroethane-d4	5.701	65	48871	4.301	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.000%
32) Benzene-d6	5.727	84	149916	3.927	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.600%
36) 1,2-Dichloropropane-d6	6.689	67	56900	4.238	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.800%
41) Toluene-d8	7.898	98	131625	3.600	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%
43) trans-1,3-Dichloroprop...	8.180	79	13567	3.070	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.400%
46) 2-Hexanone-d5	8.634	63	77130	36.776	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.560%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	49608	4.309	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.200%
66) 1,2-Dichlorobenzene-d4	12.193	152	57522	4.320	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.400%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.380	85	1684	0.149	ug/L #	81
12) 1,1-Dichloroethene	2.576	96	7302	0.887	ug/L #	54
17) Methyl tert-butyl Ether	3.361	73	3169	0.132	ug/L	95
19) 1,1-Dichloroethane	3.866	63	7257	0.384	ug/L	94
22) cis-1,2-Dichloroethene	4.663	96	35370	3.297	ug/L	99
25) Chloroform	5.087	83	47698	2.304	ug/L	96
34) Trichloroethene	6.541	95	34887	3.142	ug/L	97
47) Tetrachloroethene	8.547	164	1251	0.157	ug/L	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083123\
Data File : VU055014.D
Acq On : 31 Aug 2023 12:18
Operator : MD/SY
Sample : 04230-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH497

Quant Time: Sep 01 04:59:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 01 04:57:47 2023
Response via : Initial Calibration









#2
 Dichlorodifluoromethane
 Concen: 0.149 ug/L
 RT: 1.380 min Scan# 21
 Delta R.T. -0.003 min
 Lab File: VU055014.D
 Acq: 31 Aug 2023 12:18




Instrument :
 MSVOA_U
 ClientSampleId :
 BH497


Tgt Ion: 85 Resp: 1684
 Ion Ratio Lower Upper
 85 100
 87 42.2 25.4 38.0#



#12
 1,1-Dichloroethene
 Concen: 0.887 ug/L
 RT: 2.576 min Scan# 400
 Delta R.T. 0.000 min
 Lab File: VU055014.D
 Acq: 31 Aug 2023 12:18

Tgt Ion: 96 Resp: 7302
 Ion Ratio Lower Upper
 96 100
 61 180.8 132.4 245.8
 63 239.4 82.5 153.1#









#17
Methyl tert-butyl Ether
Concen: 0.132 ug/L
RT: 3.361 min Scan# 644
Delta R.T. 0.000 min
Lab File: VU055014.D
Acq: 31 Aug 2023 12:18




Instrument :
MSVOA_U
ClientSampleId :
BH497


Tgt Ion: 73	Resp:	3169
Ion	Ratio	Lower Upper
73	100	
43	23.2	16.2 24.2
57	20.4	17.4 26.0



#19
1,1-Dichloroethane
Concen: 0.384 ug/L
RT: 3.866 min Scan# 801
Delta R.T. 0.000 min
Lab File: VU055014.D
Acq: 31 Aug 2023 12:18

Tgt Ion: 63	Resp:	7257
Ion	Ratio	Lower Upper
63	100	
65	37.0	22.5 41.7
83	13.4	9.5 17.7





#22

cis-1,2-Dichloroethene

Concen: 3.297 ug/L

RT: 4.663 min Scan# 1108

Delta R.T. 0.000 min

Lab File: VU055014.D

Acq: 31 Aug 2023 12:18

Instrument :

MSVOA_U

ClientSampleId :

BH497





Tgt Ion: 96 Resp: 35370

Ion Ratio Lower Upper

96 100

61 125.3 88.4 164.2

98 61.0 44.0 81.6



#25

Chloroform

Concen: 2.304 ug/L

RT: 5.087 min Scan# 1181

Delta R.T. 0.003 min

Lab File: VU055014.D




Acq: 31 Aug 2023 12:18


Tgt Ion: 83 Resp: 47698

Ion Ratio Lower Upper

83 100

85 63.5 46.7 86.7









#34
Trichloroethene
Concen: 3.142 ug/L
RT: 6.541 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU055014.D
Acq: 31 Aug 2023 12:18

Instrument :
MSVOA_U
ClientSampleId :
BH497

Tgt Ion: 95	Resp: 34887
Ion Ratio	Lower Upper
95 100	
97 59.9	44.1 81.9
132 91.5	63.6 118.2
130 94.6	63.6 118.0



#47
Tetrachloroethene
Concen: 0.157 ug/L
RT: 8.547 min Scan# 2257
Delta R.T. -0.006 min
Lab File: VU055014.D
Acq: 31 Aug 2023 12:18

Tgt	Ion:164	Resp:	1251
Ion	Ratio	Lower	Upper
164	100		
129	107.0	72.8	135.2
131	75.1	70.7	131.3
166	128.3	95.9	178.1

