

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083123\
 Data File : VU055024.D
 Acq On : 31 Aug 2023 16:25
 Operator : MD/SY
 Sample : 04153-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH465

Quant Time: Sep 01 05:01:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 01 04:57:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	191556	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	190796	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	99119	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	38961	6.444	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	128.800%
7) Chloroethane-d5	1.904	69	44654	5.717	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.400%
11) 1,1-Dichloroethene-d2	2.563	65	22715	6.647	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	133.000%#
20) 2-Butanone-d5	4.628	46	125395	38.423	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.840%
24) Chloroform-d	5.059	84	99216	4.424	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.702	65	53425	4.477	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	5.724	84	198594	4.631	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	6.689	67	64813	4.297	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.000%
41) Toluene-d8	7.898	98	178454	4.345	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	8.184	79	17455	3.516	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.400%
46) 2-Hexanone-d5	8.634	63	75106	31.877	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.760%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	47519	3.674	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.400%
66) 1,2-Dichlorobenzene-d4	12.193	152	63561	4.016	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU083123\
Data File : VU055024.D
Acq On : 31 Aug 2023 16:25
Operator : MD/SY
Sample : 04153-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH465

Quant Time: Sep 01 05:01:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 01 04:57:47 2023
Response via : Initial Calibration

