

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090119\
 Data File : VU034128.D
 Acq On : 31 Aug 2019 17:13
 Operator : JC/SP
 Sample : VIBLK46
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK46

Quant Time: Sep 01 02:26:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR090119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Sep 01 01:40:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	313629	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	316188	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	148448	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	105943	5.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.20%
7) Chloroethane-d5	1.68	69	90686	5.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.00%
11) 1,1-Dichloroethene-d2	2.26	63	151408	3.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.60%
20) 2-Butanone-d5	4.18	46	256718	52.61	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.22%
24) Chloroform-d	4.62	84	177595	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.29	65	96529	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.80%
32) Benzene-d6	5.32	84	407321	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	6.31	67	126797	4.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.00%
41) Toluene-d8	7.55	98	345779	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	7.84	79	42991	4.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.40%
46) 2-Hexanone-d5	8.30	63	186566	47.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.56%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	94260	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	11.84	152	135850	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090119\
Data File : VU034128.D
Acq On : 31 Aug 2019 17:13
Operator : JC/SP
Sample : VIBLK46
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK46

Quant Time: Sep 01 02:26:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR090119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Sep 01 01:40:37 2019
Response via : Initial Calibration

