

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090121\
 Data File : VU044547.D
 Acq On : 01 Sep 2021 13:31
 Operator : SY/MD
 Sample : M3586-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PW-B6-L66-082721

Quant Time: Sep 02 11:44:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U083121DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Sep 01 01:29:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	19164	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	5598	0.763	ug/l	0.00
Spiked Amount 1.000			Recovery	=	76.000%	
68) 1,2-Dichlorobenzene-d4	12.201	152	6951	0.816	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.054	84	5133	0.362	ug/l	98
27) Chloroform	5.096	83	1195365	61.996	ug/l	99
44) Bromodichloromethane	7.115	83	74151	10.007	ug/l	100
55) Dibromochloromethane	8.819	129	7004	1.354	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090121\
Data File : VU044547.D
Acq On : 01 Sep 2021 13:31
Operator : SY/MD
Sample : M3586-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PW-B6-L66-082721

Quant Time: Sep 02 11:44:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U083121DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Sep 01 01:29:46 2021
Response via : Initial Calibration

