

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090121\
 Data File : VU044549.D
 Acq On : 01 Sep 2021 14:31
 Operator : SY/MD
 Sample : M3604-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PW-B6-L41-083021

Quant Time: Sep 02 11:44:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U083121DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Sep 01 01:29:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.121	96	18834	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	5820	0.807	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.000%	
68) 1,2-Dichlorobenzene-d4	12.198	152	6923	0.827	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.000%	
Target Compounds						
						Qvalue
9) 1,1-Dichloroethene	2.591	96	1962	0.220	ug/l	83
15) Methylene Chloride	3.057	84	3336	0.240	ug/l	97
22) cis-1,2-Dichloroethene	4.674	96	7802	0.709	ug/l	90
37) Trichloroethene	6.549	130	4950	0.840	ug/l	100
58) Tetrachloroethene	8.558	164	3377	0.608	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090121\
Data File : VU044549.D
Acq On : 01 Sep 2021 14:31
Operator : SY/MD
Sample : M3604-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PW-B6-L41-083021

Quant Time: Sep 02 11:44:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U083121DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Sep 01 01:29:46 2021
Response via : Initial Calibration

