


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090121\
 Data File : VU044541.D
 Acq On : 01 Sep 2021 08:57
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U083121DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Wed Sep 01 01:29:46 2021



AutoFind: Averaged scan 2879 to 2919; Bkg corrected with scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	2488	PASS
75	95	30	60	53.8	6780	PASS
95	95	100	100	100.0	12600	PASS
96	95	5	9	7.0	878	PASS
173	174	0.00	2	1.2	132	PASS
174	95	50	100	88.0	11088	PASS
175	174	5	9	7.8	863	PASS
176	174	95	101	96.9	10744	PASS
177	176	5	9	6.5	700	PASS