

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090122\
 Data File : VU050576.D
 Acq On : 01 Sep 2022 13:23
 Operator : SY/MD
 Sample : VU0901MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK176

Quant Time: Sep 02 02:08:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM083022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 02 02:07:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	371604	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	312428	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	114622	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	181542	63.698	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 127.400%		
7) Chloroethane-d5	1.906	69	111408	50.138	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 100.280%		
11) 1,1-Dichloroethene-d2	2.562	63	220951	45.916	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 91.840%		
21) 2-Butanone-d5	4.626	46	247030	99.106	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 99.110%		
24) Chloroform-d	5.063	84	282727	50.982	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.960%		
26) 1,2-Dichloroethane-d4	5.703	65	185989	51.632	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.260%		
32) Benzene-d6	5.726	84	608971	62.888	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 125.780%#		
36) 1,2-Dichloropropane-d6	6.690	67	214466	66.517	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 133.040%#		
41) Toluene-d8	7.896	98	497914	59.105	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 118.200%		
43) trans-1,3-Dichloroprop...	8.179	79	77810	60.006	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 120.020%		
47) 2-Hexanone-d5	8.636	63	96951	83.048	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 83.050%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	238277	47.380	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 94.760%		
66) 1,2-Dichlorobenzene-d4	12.192	152	141676	57.089	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 114.180%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090122\
Data File : VU050576.D
Acq On : 01 Sep 2022 13:23
Operator : SY/MD
Sample : VU0901MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK176

Quant Time: Sep 02 02:08:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM083022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 02 02:07:32 2022
Response via : Initial Calibration

