

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090122\
 Data File : VU050578.D
 Acq On : 01 Sep 2022 14:13
 Operator : SY/MD
 Sample : VU0901MBL02
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK178

Quant Time: Sep 02 02:08:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM083022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 02 02:07:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	291827	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	277845	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	108191	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	116440	52.024	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 104.040%			
7) Chloroethane-d5	1.903	69	82062	47.027	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 94.060%			
11) 1,1-Dichloroethene-d2	2.562	63	170926	45.231	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 90.460%			
21) 2-Butanone-d5	4.629	46	165575	84.586	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 84.590%			
24) Chloroform-d	5.064	84	234858	53.928	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 107.860%			
26) 1,2-Dichloroethane-d4	5.700	65	154348	54.561	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 109.120%			
32) Benzene-d6	5.726	84	483499	56.145	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 112.300%			
36) 1,2-Dichloropropane-d6	6.690	67	154415	53.853	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 107.700%			
41) Toluene-d8	7.899	98	409446	54.653	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 109.300%			
43) trans-1,3-Dichloroprop...	8.179	79	61704	53.508	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 107.020%			
47) 2-Hexanone-d5	8.636	63	68389	65.873	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 65.870%			
56) 1,1,2,2-Tetrachloroeth...	10.755	84	214550	47.972	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 95.940%			
66) 1,2-Dichlorobenzene-d4	12.195	152	132463	56.549	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 113.100%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090122\
Data File : VU050578.D
Acq On : 01 Sep 2022 14:13
Operator : SY/MD
Sample : VU0901MBL02
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK178

Quant Time: Sep 02 02:08:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM083022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 02 02:07:32 2022
Response via : Initial Calibration

