

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090122\
 Data File : VU050579.D
 Acq On : 01 Sep 2022 14:35
 Operator : SY/MD
 Sample : N4322-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 02 02:08:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM083022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 02 02:07:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	280787	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	270924	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	104391	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	111683	51.860	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 103.720%		
7) Chloroethane-d5	1.906	69	81127	48.319	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 96.640%		
11) 1,1-Dichloroethene-d2	2.565	63	164459	45.231	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 90.460%		
21) 2-Butanone-d5	4.626	46	167225	88.788	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 88.790%		
24) Chloroform-d	5.060	84	228597	54.554	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 109.100%		
26) 1,2-Dichloroethane-d4	5.700	65	149519	54.932	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 109.860%		
32) Benzene-d6	5.726	84	469694	55.936	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 111.880%		
36) 1,2-Dichloropropane-d6	6.690	67	150757	53.920	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 107.840%		
41) Toluene-d8	7.899	98	396228	54.239	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 108.480%		
43) trans-1,3-Dichloroprop...	8.179	79	60231	53.565	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 107.120%		
47) 2-Hexanone-d5	8.636	63	70048	69.195	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 69.190%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	214987	49.297	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 98.600%		
66) 1,2-Dichlorobenzene-d4	12.192	152	129144	57.139	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 114.280%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090122\
Data File : VU050579.D
Acq On : 01 Sep 2022 14:35
Operator : SY/MD
Sample : N4322-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 02 02:08:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM083022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 02 02:07:32 2022
Response via : Initial Calibration

