

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
 Data File : VU055061.D
 Acq On : 01 Sep 2023 23:57
 Operator : MD/SY
 Sample : 04230-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Sep 02 00:24:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 02 00:23:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	184160	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	185363	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	94138	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	41520	4.810	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.200%
7) Chloroethane-d5	1.911	69	47849	4.789	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.800%
11) 1,1-Dichloroethene-d2	2.563	65	24772	5.037	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.800%
20) 2-Butanone-d5	4.628	46	145824	58.220	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.440%
24) Chloroform-d	5.058	84	117300	5.121	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
26) 1,2-Dichloroethane-d4	5.698	65	60486	5.279	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.724	84	231343	5.056	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.200%
36) 1,2-Dichloropropane-d6	6.689	67	75485	5.187	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.800%
41) Toluene-d8	7.898	98	205513	4.876	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
43) trans-1,3-Dichloroprop...	8.181	79	19376	4.835	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.634	63	85638	52.634	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.260%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	56133	5.485	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	74573	5.427	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.600%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.576	96	17061	2.004	ug/L #	43
19) 1,1-Dichloroethane	3.872	63	7462	0.400	ug/L #	88
22) cis-1,2-Dichloroethene	4.663	96	63833	6.037	ug/L	97
25) Chloroform	5.087	83	13179	0.678	ug/L	100
27) 1,2-Dichloroethane	5.795	62	2227	0.188	ug/L #	76
34) Trichloroethene	6.541	95	59155	5.395	ug/L	97
35) Methylcyclohexane	6.685	83	17482	1.136	ug/L #	20
37) 1,2-Dichloropropane	6.689	63	8881	0.827	ug/L #	88
42) Toluene	7.968	91	14345	0.341	ug/L	98
48) 2-Hexanone	8.682	43	20336	4.238	ug/L #	88
61) 1,2,3-Trichloropropane	11.807	75	14223	2.277	ug/L #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090123\
Data File : VU055061.D
Acq On : 01 Sep 2023 23:57
Operator : MD/SY
Sample : 04230-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :


Quant Time: Sep 02 00:24:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 02 00:23:04 2023
Response via : Initial Calibration








#12
 1,1-Dichloroethene
 Concen: 2.004 ug/L
 RT: 2.576 min Scan# 400
 Delta R.T. 0.000 min
 Lab File: VU055061.D
 Acq: 01 Sep 2023 23:57

Instrument :
 MSVOA_U
 ClientSampleId :







Tgt Ion: 96 Resp: 17061
 Ion Ratio Lower Upper
 96 100
 61 186.8 120.8 224.4
 63 273.5 94.5 175.5#



#19
 1,1-Dichloroethane
 Concen: 0.400 ug/L
 RT: 3.872 min Scan# 803
 Delta R.T. 0.006 min
 Lab File: VU055061.D
 Acq: 01 Sep 2023 23:57

Tgt Ion: 63 Resp: 7462
 Ion Ratio Lower Upper
 63 100
 65 31.6 22.2 41.2
 83 0.0 10.6 19.8#









#22
 cis-1,2-Dichloroethene
 Concen: 6.037 ug/L
 RT: 4.663 min Scan# 1104
 Delta R.T. 0.000 min
 Lab File: VU055061.D
 Acq: 01 Sep 2023 23:57




Instrument :
 MSVOA_U
 ClientSampleId :


Tgt Ion: 96 Resp: 63833
 Ion Ratio Lower Upper
 96 100
 61 133.4 97.2 180.4
 98 65.4 46.1 85.7



#25
 Chloroform
 Concen: 0.678 ug/L
 RT: 5.087 min Scan# 1181
 Delta R.T. 0.003 min
 Lab File: VU055061.D
 Acq: 01 Sep 2023 23:57

Tgt Ion: 83 Resp: 13179
 Ion Ratio Lower Upper
 83 100
 85 67.1 47.2 87.6









#27
1,2-Dichloroethane
Concen: 0.188 ug/L
RT: 5.795 min Scan# 1401
Delta R.T. 0.003 min
Lab File: VU055061.D
Acq: 01 Sep 2023 23:57




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 62 Resp: 2227
Ion Ratio Lower Upper
62 100
98 0.0 6.9 10.3#



#34
Trichloroethene
Concen: 5.395 ug/L
RT: 6.541 min Scan# 1633
Delta R.T. 0.000 min
Lab File: VU055061.D
Acq: 01 Sep 2023 23:57

Tgt Ion: 95 Resp: 59155
Ion Ratio Lower Upper
95 100
97 61.6 44.7 83.1
132 90.7 66.6 123.6
130 93.4 66.5 123.5





#35

Methylcyclohexane

Concen: 1.136 ug/L

RT: 6.685 min Scan# 1

Delta R.T. -0.074 min

Lab File: VU055061.D

Acq: 01 Sep 2023 23:57

Instrument :

MSVOA_U

ClientSampleId :





Tgt Ion: 83 Resp: 17482

Ion Ratio Lower Upper

83 100

55 0.3 60.4 90.6#

98 1.1 35.3 52.9#



#37

1,2-Dichloropropane

Concen: 0.827 ug/L

RT: 6.689 min Scan# 1679

Delta R.T. -0.096 min

Lab File: VU055061.D




Acq: 01 Sep 2023 23:57


Tgt Ion: 63 Resp: 8881

Ion Ratio Lower Upper

63 100

112 0.0 3.2 4.8#





#42

Toluene

Concen: 0.341 ug/L

RT: 7.968 min Scan# 2076

Delta R.T. 0.003 min

Lab File: VU055061.D

Acq: 01 Sep 2023 23:57

Instrument :

MSVOA_U





ClientSampleId :

Tgt Ion: 91 Resp: 14345

Ion Ratio Lower Upper

91 100

92 61.0 41.9 77.9



#48

2-Hexanone

Concen: 4.238 ug/L

RT: 8.682 min Scan# 2299

Delta R.T. 0.000 min

Lab File: VU055061.D

Acq: 01 Sep 2023 23:57

Tgt Ion: 43 Resp: 20336




Ion Ratio Lower Upper


43 100

58 43.6 44.0 66.0#

57 19.7 16.2 24.2


100 6.3 9.0 13.4#





#61
 1,2,3-Trichloropropane
 Concen: 2.277 ug/L
 RT: 11.807 min Scan# 31
 Delta R.T. 0.987 min
 Lab File: VU055061.D
 Acq: 01 Sep 2023 23:57

Instrument :
 MSVOA_U
 ClientSampleId :



Tgt Ion:	75	Resp:	14223
Ion	Ratio	Lower	Upper
75	100		
110	2.5	28.2	42.2#
77	30.8	26.2	39.4

